

Original Research

Optimized Simultaneous ASL and BOLD Functional Imaging of the Whole Brain

Vincent J. Schmithorst, PhD,^{1,2} Luis Hernandez-Garcia, PhD,³ Jennifer Vannest, PhD,^{1*} Akila Rajagopal, MS,¹ Greg Lee, PhD,¹ and Scott K. Holland, PhD¹

Purpose: To compare a double-excitation combined arterial-spin labeling/blood-oxygenation level dependent (ASL/BOLD) functional imaging method to a double-echo method. ASL provides a useful complement to standard BOLD functional imaging, to map effects of cerebral hemodynamics. Whole-brain imaging is necessary to properly characterize large functional networks. A challenge of whole-brain ASL/BOLD is that images for ASL functional contrast must be acquired before significant longitudinal relaxation of the inverted spins occurs; however, a longer echo time (TE) is required for optimal BOLD functional contrast, lengthening the acquisition time. Thus, existing combined ASL/BOLD studies have only partial-brain coverage.

Materials and Methods: The proposed method allows acquisition of images for ASL contrast within a short period after the ASL labeling pulse and postinversion delay, then subsequent acquisition of images with longer TE for BOLD contrast. The technique is demonstrated using a narrative comprehension task in 35 normal children, and the double-excitation method is empirically compared with the double-echo method in 7 normal adults.

Results: Compared with a double-echo sequence, simulations show the double-excitation method improves ASL contrast-to-noise ratio (CNR) (~50%) in later-acquired slices with minimal (<1%) reduction in BOLD CNR in earlier-acquired slices if reduced excitation flip angles for the ASL acquisitions are used. Empirical results from adult data are in agreement with the simulations. Group analyses from the narrative comprehension task also show greater intersubject sensitivity in BOLD versus ASL.

Conclusion: Our method simultaneously optimizes ASL and BOLD acquisitions for CNR while economizing acquisition time.

Key Words: arterial spin labeling; BOLD functional imaging; pediatric neuroimaging

J. Magn. Reson. Imaging 2014;39:1104–1117.

© 2013 Wiley Periodicals, Inc.

OBSERVED CHANGES IN BOLD activity associated with age during development are the result of developmental changes in hemodynamics as well as neuronal activity (1). This confound is of particular concern for functional MRI studies of brain development, where neurocognitive changes may occur simultaneously with neurophysiological changes in cerebral perfusion. Indeed, this issue could influence future research involving the “child connectome,” the developing structural and functional connectivity underlying the development of higher-order cognitive function (2). Developmental changes in hemodynamics must be dissociated from neurocognitive development to properly interpret data from intrinsic-connectivity (resting-state) as well as task-based BOLD fMRI studies investigating functional connectivity.

A possible solution could be a “calibrated fMRI” technique, involving either a hypercapnic or hyperoxic challenge (3), which allows a direct estimate of cerebral metabolic rate of oxygen (CMRO₂) by accounting for signal effects due to changes in blood flow. However, such measurements may not be practical or feasible, especially in pediatric populations. Alternatively, one could rely exclusively on functional contrast or connectivity found from arterial spin labeling (ASL) techniques, which focus on cerebral perfusion, because perfusion has been shown to be linearly related to CMRO₂ (4). However, the coupling constant between blood flow and CMRO₂ likely varies significantly over the developmental period, making developmental inferences difficult. Moreover, functional ASL provides significantly less intrasubject CNR than standard BOLD (5,6) unless the stimulation paradigm is very long. Moreover, the shortest repetition time (TR) possible with ASL is more than twice as long as in standard BOLD, because of the additional time

¹Cincinnati Children’s Hospital Medical Center, Pediatric Neuroimaging Research Consortium, Cincinnati, Ohio, USA.

²Children’s Hospital of Pittsburgh of UPMC, Dept. of Radiology, Pittsburgh, Pennsylvania, USA.

³University of Michigan, Functional MRI Laboratory, 1096 BIRB, Ann Arbor, Michigan, USA.

Contract grant sponsor: National Institute of Child Health and Human Development (NIH); Contract grant number: HHSN275200900018C; Contract grant title: The Pediatric Functional Imaging Research Network.

*Address reprint requests to: J.V., Cincinnati Children’s Hospital Medical Center, Pediatric Neuroimaging Research Consortium, 3333 Burnet Avenue, Cincinnati, OH 45229. E-mail: jennifer.vannest@cchmc.org

Received October 15, 2012; Accepted May 16, 2013.

DOI 10.1002/jmri.24273

View this article online at wileyonlinelibrary.com.




Figure 1. Diagram of the proposed double-excitation sequence for simultaneous ASL/BOLD imaging of brain activity. The gray-shaded ASL labeling block indicates a control label that produces no inversion but identical magnetization transfer effects as the pCASL inversion block.

allocated to labeling and the need to obtain both labeled and control images.

Alternatively, one could acquire ASL perfusion contrast and BOLD contrast simultaneously. These acquisitions can be designed as single-excitation single-echo, single-excitation double-echo, or double-excitation, depending on the number of excitation pulses per slice and the number of echoes per slice following the ASL labeling pulse. In the single-excitation single-echo approach, the same acquisition is used both for BOLD and ASL contrast. This results in a tradeoff between BOLD and ASL contrast depending on the echo time (TE). Short TE, optimal for ASL acquisitions, is sub-optimal for BOLD contrast but using a longer TE to increase BOLD weighting, reduces the ASL contrast-to-noise (CNR) due to increased attenuation of the baseline signal. Double-echo techniques (7–9) are, therefore, an attractive alternative. This approach allows a short-TE acquisition for ASL contrast (first echo) followed by a second echo with a TE optimized for BOLD contrast. With this method, however, each slice takes twice as long to acquire. Because of the additional time spent per slice, the method suffers from reduced ASL CNR for later-acquired slices as compared to the single-echo approach, because the labeled spins will have had more time for longitudinal relaxation. As a result, the number of slices one can acquire per time point before the label clears out from the brain is reduced by half (the single-echo single-excitation approach with a longer TE also suffers from this limitation, although to a lesser extent).

A different approach is a double-excitation strategy (Fig. 1), which allows for simultaneous optimization of both ASL and BOLD CNR. The first excitation is used to collect the ASL slices at a short TE, and the second excitation occurs immediately afterward to collect the BOLD slices at a longer TE. In this configuration, the ASL slices can be acquired as fast as possible after the labeling pulse and inflow delay to maximize contrast from the label. This approach has been used in a few previous studies (e.g., (10–12)), although none of them permitted whole-brain coverage. Kastrup et al (10) acquired a single slice using a flow-sensitive alternating inversion recovery (FAIR) labeling

technique. St Lawrence et al (12) used a pulsed labeling technique with attenuating the static signal in arterial spin tagging (ASSIST) background suppression (13). This strategy necessitated a long 180-degree pulse in between clusters of slices for the ASL acquisitions to maintain suppression of the background signal, again reducing the CNR available at the later-acquired slices. Stefanovic et al (11) also used a similar technique, limiting the number of slices acquired to 6.

The limited number of slices and, therefore, brain coverage, is of particular importance for studies involving brain connectivity mapping. Both task-based and intrinsic-connectivity fMRI studies have revealed networks covering the entire reach of the brain, even from the earliest periods in development (14–18). Therefore, while partial-brain coverage is useful for elucidating local relationships between neuronal function and hemodynamics (e.g., Stefanovic et al) (11) whole-brain coverage is essential to properly characterize the development and complex interplay of functional brain networks. However, until now the considerations outlined above have limited combined ASL/BOLD imaging to partial brain coverage (typically between 6 and 12 slices are acquired) rendering it a novel curiosity for a few specialists but limiting its value in brain connectivity mapping. Here, we introduce a new method for simultaneous ASL/BOLD functional imaging that yields full brain coverage with adequate spatial and temporal resolution for whole brain mapping. The method we have implemented uses standard hardware configuration delivered by the manufacturer with radiofrequency (RF) body coil for transmission and 32-channel head coil for echo detection with SENSE reconstruction. Limitations related to a pulsed labeling strategy are overcome by using either continuous (CASL) or pseudo-continuous (pCASL) labeling. pCASL labeling may be performed using the body coil without the need for separate hardware (e.g., a separate neck coil) as used for CASL.

A possible drawback of the double-excitation strategy could be reduced CNR available for the BOLD acquisitions because of partial saturation from the excitation pulses of the preceding ASL acquisition.

Table 1
Parameter values used for the simulations^a

δ_a (arterial transit time)	500 ms (24)
δ (tissue transit time)	1500 ms (23)
τ (labeling duration)	1500 ms (typical value in our laboratory)
w (post-inversion delay)	700 ms (21)
CBF (baseline)	60 ml/(100g min) (20)
CBF (increase over baseline)	50% (20)
BOLD signal (increase over baseline)	2% at TE = 35 ms (typical value in our laboratory) Corresponds to $\Delta R2^* = 0.566 \text{ s}^{-1}$
λ (blood-brain partition coefficient)	0.9 ml blood/g brain
α (labeling efficiency)	0.8 (Typical value in our laboratory)
$T1_a$ (T1 of arterial blood)	1660 ms (22)
$T1_{app}$ (apparent T1 in gray matter)	1300 ms (43)
$T2^*$ of gray matter at 3 Tesla	40 ms (43)
# slices	30
TE	Double-excitation: 11 ms for ASL, 35 ms for BOLD Double-echo: 11 ms for ASL, 35 ms for BOLD Single-excitation single-echo: 15 ms, 20 ms, 25 ms, 30 ms
Time to acquire each slice	TE + 7 ms
TR	Double-excitation: 3958 ms Double-echo: 3418 ms Single-excitation single-echo: 2838 ms, 2983 ms 3128 ms, 3273 ms
SNR_0 (Image SNR at TE = 30 ms)	866 (19)
λ (physiological noise constant at TE = 30–35 ms)	0.0092 (19,25)
λ (physiological noise constant at TE = 25 ms)	0.0083 (25)
λ (physiological noise constant at TE = 20 ms)	0.0075 (25)
λ (physiological noise constant at TE = 15 ms)	0.0067 (25)
λ (physiological noise constant at TE = 10–11 ms)	0.0059 (25)

^aReferences are given for parameters taken from the literature.

However, functional ASL CNR is lower overall; thus, it is advantageous to trade off BOLD CNR for ASL CNR. Moreover, for typical voxel sizes (e.g., $4 \times 4 \times 4$ mm) physiological noise dominates over thermal noise (19), which will lessen the overall CNR reduction from the reduced effective TR (in this study, we use the term “effective TR” to refer to the period of time available for recovery of longitudinal magnetization, and “actual TR” to refer to the actual repetition time). Following the approach of Gonzalez-Castillo et al (19), which demonstrated minimal CNR reduction in BOLD acquisitions for excitation flip angles much below the Ernst angle, we also investigate the potential of lessening the BOLD CNR reduction by reducing the excitation flip angles for the ASL acquisitions. In a multislice sequence, the time between the ASL and BOLD acquisitions and hence the effective TR will be shorter in the first-

acquired slices, because the BOLD acquisitions have a longer TE and, therefore, longer slice acquisition time.

METHODS

ASL/BOLD Theory and Simulations

We simulated the ASL and BOLD signals obtained using three different methods: (i) the proposed (double-excitation) technique, (ii) a single-excitation single-echo acquisition, and (iii) a double-echo acquisition. This allows comparison between the SNR and CNR of the proposed acquisition scheme in relation to the other two. Simulations using a continuous ASL (CASL) sequence were performed using IDL (Exelis, Boulder, CO). The signal change due to application of the CASL tagging pulse was estimated using a theoretical model derived from a general kinetic model (20) as:

$$\frac{\Delta M}{M_0} = \frac{-2\alpha f}{\lambda R_{1app}} e^{-\delta R_{1a}} (e^{R_{1app} \min(\delta-w, 0)} - e^{R_{1app}(\delta-\tau-w)}) - \left. \begin{array}{l} \\ \frac{2\alpha f}{\lambda R_{1a}} (e^{-R_{1a}\delta_a} e^{R_{1a} \min(0, -w+\delta_a)} - e^{-R_{1a}\delta} e^{R_{1a} \min(0, -w+\delta)}) \end{array} \right\}, \tau + w > \delta$$

$$\frac{\Delta M}{M_0} = \frac{-2\alpha f}{\lambda R_{1a}} (e^{R_{1a}(\min(\delta_a-w, 0)-\delta_a)} - e^{-R_{1a}(\tau+w)}), t + w < \delta, \quad [1]$$

where $\frac{\Delta M}{M_0}$ = the signal change; λ = blood-brain partition coefficient; f = cerebral blood flow; w = postinversion delay; τ = labeling duration; R_{1a} = arterial R1; α = labeling efficiency, R_{1app} = apparent tissue R1; δ = tissue transit time; δ_a = arterial transit time. This model does not assume all spins have flowed into the voxel or that all spins have traversed from the arterial compartment into the tissue compartment. This model, however, does assume a single transit time for all spins; differences in signal change between a single transit time model and a distributed transit time model have been shown to be negligible (21).

The temporal SNR (tSNR) is a function both of physiological and thermal noise (19). Thermal noise is independent of MR signal strength, while physiological noise is proportional to MR signal strength. tSNR is related to image SNR (19) by:

$$tSNR = \frac{SNR}{\sqrt{1 + \lambda^2 SNR^2}} \quad [2]$$

where λ is the proportionality constant between physiological noise (σ_p) and MR signal strength (SS): $\sigma_p = \lambda SS$. Image SNR is only a function of thermal noise and will be linearly related to MR signal strength: $SNR = SNR_0 \frac{SS}{SS_0}$ where SNR_0 is the SNR for a baseline signal strength (SS_0) and SS is the actual signal strength, which will be attenuated from SS_0 by factors such as varying flip angle or varying effective TR.

Parameter values chosen for the simulation are either typical literature values or typical values used for implementation of the sequence at 3 Tesla (T) and are listed in Table 1 (22–24). For the ASL acquisitions, labeling efficiencies of 0.8 and above are routinely achieved on a Philips Achieva 3T system using a pCASL labeling technique. A value of 700 ms for post-inversion delay was chosen (21) as a compromise

between maximizing ASL functional contrast due to increased CBF while minimizing spurious contrast due to decreased transit time during functional activation. As the approximate time of peak ASL CNR occurs when the postinversion delay is equal to the arterial transit time, using this value is only slightly suboptimal for the first slices.

For our simulated double-excitation or double-echo scans, we used the following parameters as typical values for use of the sequence at 3T: $TE_{ASL} = 11$ ms; $TE_{BOLD} = 35$ ms; TA (time to acquire a slice) = TE + 7 ms; # slices acquired = 30; TR = minimum, $T2^* = 40$ ms. For the simulated single-excitation single-echo scan, the echo time was allowed to vary; we used echo times of 15 ms, 20 ms, 25 ms, and 30 ms.

For the estimation of BOLD tSNR, we used published values (19) of $\lambda = .0092$, $SNR_0 = 866$ (resulting in tSNR₀ of approximately 100) for gray matter at 3T for an echo time of 30 ms and voxel resolution of 3.75

$\times 3.75 \times 4$ mm. For the estimation of ASL tSNR (with a shorter TE), we used the data of Kruger and Glover (25) to estimate an appropriate value for λ (which is TE-dependent). The physiological noise σ_P is expressed as $\sigma_P = \sqrt{\sigma_B^2 + \sigma_{NB}^2}$ where σ_B is related to the physiological fluctuation in R_2^* and is thus TE dependent, while σ_{NB} is related to contributions with no TE dependency. Using the published values of:

$$\frac{\sigma_B(TE = 10ms)}{\sigma_B(TE = 30ms)} \cong 0.44, \sigma_B(TE = 30ms) = 0.53, \sigma_{NB} = 0.21, \sigma_0 = 0.14, S = 52.9, \quad [3]$$

where σ_0 is the thermal noise and S is the mean signal, we computed $\sigma_P = 0.314$ and $\lambda = \sigma_P/S = 0.0059$. For the single-excitation single-echo sequences, as σ_B varies approximately linearly in the range from TE = 10 ms to TE = 30 ms, the value of σ_B was estimated by means of linear interpolation and λ computed as above (values listed in Table 1).

Image SNR was adjusted from SNR_0 for the effective TR (e.g., available relaxation time from previous excitation pulse) and for the difference in TE (SNR_0 was given for TE = 30 ms). tSNR was calculated as in Eq. [2], using the appropriate values of image SNR and λ .

To estimate BOLD CNR, we used a typical value at 3T for BOLD signal increase over baseline of 2% at TE = 35 ms and thus BOLD CNR was estimated as:

$$CNR_{BOLD} = \frac{(BOLD_{functional} - BOLD_{baseline}) * tSNR}{BOLD_{baseline}} = BOLD_{CONTRAST} * tSNR, \quad [4]$$

where for the single-excitation single-echo sequences which use shorter TEs, BOLD_{CONTRAST} was adjusted from the 2% value used at TE = 35 ms.

To estimate ASL CNR, the following procedure was used. The functional contrast in ASL is given by:

$$\Delta M_{functional,ASL} = (fASL_{control} - fASL_{labeled}) - (bASL_{control} - bASL_{labeled}), \quad (5)$$

where fASL and bASL are the ASL acquisitions occurring during the functional and baseline conditions,

respectively. The CNRs of the functional and baseline conditions are:

$$CNR_{functional} = \frac{(fASL_{control} - fASL_{labeled})}{fASL_{control}} * tSNR, \quad [6]$$

$$CNR_{baseline} = \frac{(bASL_{control} - bASL_{labeled})}{bASL_{control}} * tSNR$$

where tSNR represents the temporal SNR of the ASL images (assuming this does not vary significantly between baseline and functional acquisitions).

The fractional signal difference between label and control acquisitions is given by Eq. [1] and is linearly proportional to CBF. For the difference between the functional and the baseline condition, we used a typical value of CBF increase of 50% over baseline, and thus:

$$\Delta M_{functional,ASL} = (fASL_{control} - fASL_{labeled}) - (bASL_{control} - bASL_{labeled}), \quad [7]$$

$$CNR_{functional} \cong 1.5 * \frac{(bASL_{control} - bASL_{labeled})}{bASL_{control}} * tSNR, \quad [8]$$

The ASL CNR is, therefore, given by:

$$CNR_{ASL} = 0.5 * \frac{(bASL_{control} - bASL_{labeled}) * tSNR_{ASL}}{\sqrt{2} bASL_{control}} = \frac{1}{2\sqrt{2}} * \frac{\Delta M}{M_0} * tSNR, \quad [9]$$

where $\Delta M/M_0$ is given from Eq. [1], using the CBF value for the baseline condition (the factor of square root of 2 in the denominator is due to the overall contrast being the difference between functional and baseline).

We computed the BOLD CNR and the ASL CNR for the double-excitation, the double-echo, and the single-excitation single-echo sequences for the parameters above as a function of slice number. For the single-excitation single-echo sequence, we varied the TE, using values used of 15 ms, 20 ms, 25 ms, and 30 ms. We also computed maximum CNR per unit square root of time, as the different strategies allow for different minimum TRs (listed in Table 1).

Additional Simulations

For the double-excitation sequence, we also investigated the effect of reducing the excitation flip angle for the ASL acquisitions; we varied the flip angle from 90 degrees to 30 degrees in increments of 10 degrees. We also simulated the effect of physiological noise reduction by means of techniques such as cardiac and respiratory cycle monitoring (26), assuming that such techniques are capable of reducing physiological noise to approximately 40% of its precorrection magnitude.

ASL/BOLD fMRI Experiments in Human Subjects

Narrative Comprehension Task in Children

The double-excitation technique described above was implemented on a Philips 3T Achieva system. Gradient-echo echo planar imaging (EPI) imaging

parameters were: $TE_{ASL} = 11$ ms; $TE_{BOLD} = 32$ ms; # slices acquired = 24; TR = 4000 ms; EPI matrix = 64×64 ; in-plane resolution = 3.75 mm \times 3.75 mm; slice thickness = 5 mm. A balanced pCASL sequence was used with: label duration = 1500 ms; postinversion delay = 700 ms; mean B1 = 1 μ T; pulse duration = 500 μ s; time between pulses = 1500 μ s; max gradient = 0.6 mT/m; mean gradient = 0.1 mT/m. For this study, the excitation flip angle for the ASL acquisitions was 90 degrees.

Participants were 35 typically developing healthy children ages 5–18 years performing a narrative comprehension task. The study was approved by the local Institutional Review Board (IRB); informed consent and assent (where appropriate) were obtained from a parent/guardian and the participant, respectively.

The narrative comprehension task uses a 64-s alternating block design in which a different story, read by an adult female speaker, is presented during each active block. Each story was designed by a speech-language pathologist to be appropriate for young children. During the control blocks, to control for sublexical auditory processing, broadband noise in the frequency range of speech is presented in a gradual sweep across frequencies (500–4000 Hz). The duration of each sweep varies from 0.5 to 4 s. Five 64-s active blocks and five 64-s control blocks are presented in an alternating sequence for a total scan time of 10 min 40 s. This comparison between naturalistic speech and nonspeech auditory processing is designed to engage multiple aspects of language processing while controlling for purely auditory stimulation.

Data were postprocessed using routines written in IDL (Exelis, Boulder, CO) incorporated into the LONI postprocessing pipeline (UCLA) (27,28). The ASL (short TE) and BOLD (long TE) images were separated and postprocessed separately. Data was motion-corrected using a pyramid iterative algorithm (29), sequentially using each of the acquired volumes as the reference. Using an intensity-based cost function (30) as the metric of comparison the optimal reference volume for motion correction was then selected. After motion correction, data were transformed into the MNI space using a nonlinear transformation derived from the T1-weighted anatomical image, a pediatric template (31) and the coregistration routine available in SPM8 (Wellcome Dept. of Cognitive Neurology, London, UK). Data were spatially filtered using a Gaussian filter with $\sigma = 7$ mm. BOLD functional activation was computed using a general linear model (GLM) with design matrix including task condition as the regressor of interest; linear and quadratic drift terms, parameters from the motion correction, and whether the image was acquired during the label or control pCASL tagging condition were included as covariates of no interest. For the ASL data, as is typically done, surround subtraction (32,33) was performed before the GLM on both the data and the design matrix. ASL functional activation was computed using a GLM and a similar design matrix as for the BOLD analysis but with a pCASL tagging condition by task interaction as the regressor of interest and no drift terms included as covariates.

For the second-level (group) analysis, a one-sample t-test was used. A Monte Carlo simulation was used to find an intensity ($t \geq 3.25$) and spatial-extent threshold (50 contiguous voxels) corresponding to a significance value of $P < 0.01$ with family-wise error correction. We made a comparison of overall within-group sensitivity between the ASL and BOLD images acquired using our double excitation experiment. The mean Z-score for all voxels with significant group activation for both the BOLD analysis and the ASL analysis was computed for all participants. To account for the expected greater intrasubject BOLD sensitivity (5,6) leading to higher Z-scores for the BOLD analysis, we normalized the Z-scores for each analysis to unity mean and compared the variances, after regressing out age and sex.

Estimation of Physiological Noise Parameter at TE = 11 ms

A slight discrepancy between the physiological noise parameter at TE = 11 ms estimated by Kruger and Glover (25) and our sequence as actually implemented is expected for two reasons (34). First, we used a 32-channel coil, as opposed to a single channel transmit-receive. In addition we used parallel imaging with an acceleration factor of 2 (which is necessary to reduce the total readout time to achieve a TE as short as 11 ms). However, Triantafyllou et al (34) have shown minimal dependence of λ on number of coil elements, and on parallel imaging with a low (2) acceleration factor at the lower spatial resolution we are using here. To validate the choice of physiological noise parameter λ at TE = 11 ms empirically, we used baseline CBF data available from ASL measurements in the same 30 children at rest. These are the same participants who also performed the narrative processing task described above during ASL/BOLD fMRI. The resting state ASL data can be used to compute CBF maps, including pairs of label-control acquisitions (with no functional task). T1 maps were also computed in the same subjects based on an inversion-recovery EPI sequence with the same voxel size, acquired during the same session as the CBF and ASL/BOLD image data.

Scan parameters for the baseline CBF scans were identical to the narrative processing task with the exception that there was no BOLD acquisition following the ASL acquisitions, and the postinversion delay was lengthened to 1500 ms. Scan parameters for the T1 maps were also identical with the exception that the ASL labeling was omitted, and inversion-recovery (180-degree) pulses preceded each slice: TI values used were 100, 200, 300, 500, 750, 1000, 2000, and 3000 ms. T1 maps were constructed by fitting the inversion-recovery curve using routines written in IDL.

For estimation of λ , we used the technique similar to that of Gonzalez-Castillo et al (19), detailed as follows: (i) motion-correction of the CBF data, using the technique detailed in the previous section; (ii) visual inspection of the data for residual motion artifacts which would artifactually reduce the calculated tSNR

and inflate the value of λ : two datasets with apparent visual motion were rejected; (iii) restricting analysis to voxels in gray matter by using only voxels with $1.2 \text{ s} < T_1 < 1.5 \text{ s}$; (iv) estimation of image noise by selection of an ROI far away from the head, computing the spatial standard deviation, and multiplying by a factor of $\sqrt{\frac{2}{4-\pi}} = 1.5264$ (35) to account for the non-Gaussianity of the magnitude signal; noise estimates were averaged across all volumes; (v) estimate of average image signal by averaging the signal across all gray matter voxels and across all acquired volumes, computed separately for labeled and control acquisitions; (vi) computation of spatial SNR as mean image signal/image noise (i.e., (5)/(4)); (vii) estimation of temporal SNR for each voxel, after linear and quadratic de-trending, as signal mean divided by standard deviation, computed separately for labeled and control acquisitions; (viii) computation of tSNR as the mean over all voxels and over all acquisitions (labeled or control); (ix) computation of λ by solving Eq. [2].

In SENSE reconstruction (unlike sum-of-squares reconstruction without parallel acceleration), the real and imaginary data from each coil are linearly summed separately (with weighting factors from the unfolding matrix) and afterward the image magnitude is computed. Therefore, estimation of SNR by means of use of background signal, as we are doing in step 4), is subject to some bias and should not be performed if the goal of such measurements is to compare coils or pulse sequences (36). However, the amount of such bias is not very large for the technique of estimating spatial standard deviation, which overestimates SNR by approximately only 17%, as found empirically (36); thus, we deem it sufficiently accurate for our purposes of estimating the physiological noise constant λ , which is not sensitive to slight misestimation of SNR when $\text{SNR} \gg \text{tSNR}$.

Empirical Comparison of Double-Excitation Versus Double-Echo Sequence

The narrative comprehension paradigm was also performed by a cohort of seven normal adults (one M, six F, age = 39.3 ± 13.6 years). IRB approval was also obtained for this study and informed consent obtained from all participants. The paradigm was also

implemented using the double-echo acquisition technique with the same TEs and TR as used for the double-excitation approach. For the double-excitation approach, the excitation flip angle used for the ASL acquisitions was 60 degrees. Technique order was counterbalanced across participants.

fMRI data processing was carried out using FSL (FMRIB's Software Library, www.fmrib.ox.ac.uk/fsl). Data were first motion corrected using MCFLIRT (37). The time-series were high pass filtered at a period of 128 s and 6 mm isotropic spatial smoothing was applied. The general linear model was set up for use with unsubtracted ASL data using the approach described by Mumford et al (38). This approach was used to enable a more accurate comparison between BOLD and ASL CNR. There are three primary regressors in the design matrix: (i) An alternating control/tag regressor to represent the baseline ASL signal, (ii) A BOLD task regressor created by convolving the block design with a gamma variate hemodynamic response function, and (iii) An ASL task regressor formed as the modulation of regressors 1 and 2. The six motion correction time-courses output by MCFLIRT were included as nuisance regressors in the design. Time-series statistical analysis was carried out using FILM with local autocorrelation correction (39). Affine registration of the functional volumes to the subject's T1 structural image was followed by a nonlinear warping to MNI space using symmetric diffeomorphic image registration as implemented in the Advanced Normalization Tools (ANTs) software package (40).

Empirical comparison between the double-excitation and the double-echo sequence was performed as follows. A fixed-effects analysis was performed to define ROIs for each technique (double-excitation versus double-echo). Regions were deemed significant at $T > 4$, 100 contiguous voxels for ASL; $T > 8$, 100 contiguous voxels for BOLD. The ROIs were defined as the union of all voxels meeting significance for either technique. The average T-score in the ROI was computed for each participant. Results were compared using a paired t-test; in addition, the percentage loss in CNR of the worse-performing technique was estimated as:

$$\text{CNR}_{\text{loss}} = \begin{cases} \frac{\bar{T}_{\text{double-excitation}} - \bar{T}_{\text{double-echo}}}{\bar{T}_{\text{double-excitation}}} * 100, \bar{T}_{\text{double-excitation}} > \bar{T}_{\text{double-echo}} \\ \frac{\bar{T}_{\text{double-echo}} - \bar{T}_{\text{double-excitation}}}{\bar{T}_{\text{double-echo}}} * 100, \bar{T}_{\text{double-echo}} > \bar{T}_{\text{double-excitation}} \end{cases} \quad [10]$$

RESULTS

Theoretical Simulations

Comparisons of the double-excitation versus the double-echo and single-excitation single-echo techniques for ASL CNR and CNR per unit time as a function of slice number are plotted in Figure 2. Our results clearly show that the advantage in double-excitation approach is that it preserves the ASL CNR

remaining in the brain in the later slices much better than the other acquisition schemes. At the 24th slice, the ASL CNR obtained using the double-excitation method shows an almost 50% improvement over that obtained using the double-excitation method (Fig. 2, top). The double-excitation method also out-performs all the single-excitation single-echo methods. Accounting for the differences in minimum TR between the sequences (Fig. 2, bottom), the double-




Figure 2. Diagram of ASL CNR (top) and CNR per unit time (bottom) at TR = minimum as a function of slice number for the proposed double-excitation acquisition, a double-echo acquisition, and single-excitation single-echo acquisitions at varying TEs.

excitation sequence outperforms all the other sequences except for the single-excitation single-echo with TE = 15 ms (for which performance is comparable), and the first five slices in the double-echo sequence. However, the ASL CNR per unit time still shows a 40% improvement over the double-echo method at the 24th slice, and only suffers approximately a 7% reduction in CNR per unit time for the first slice acquired.

Comparisons of the double-excitation versus the double-echo and single-excitation single-echo techniques for BOLD CNR and CNR per unit time are plotted in Figure 3 as a function of slice number. In the earlier slices, the double-excitation sequence suffers some CNR loss (~7%) as a result of the reduced effective TR (which is more pronounced in the earlier-acquired slices) as compared to the double-echo sequence (Fig. 3, top). Both the double-excitation and the double-echo sequences clearly outperform the single-excitation single-echo sequences due to the longer TE. Accounting for the differences in minimum TR between the sequences (Fig. 3, bottom), the double-excitation sequence suffers an approximately 13% reduction in CNR per unit time (in the earliest slices) compared with the double-echo sequence, but overall outperforms all the single-excitation single-echo sequences.

The effect of reducing the ASL excitation flip angle is plotted in Figure 4. A minimal reduction in ASL CNR is seen even for reducing the flip angle to 30 degrees (Fig. 4, top; intermediate values of flip angle not plotted), due to the very weak dependence of CNR

on flip angle in the regime where physiological noise dominates (19). However, approximately a 10% relative improvement in BOLD CNR is available (Fig. 4, bottom) when the flip angle is reduced to 60 degrees, with minimal improvements obtained from reducing the flip angle further.

The effect of physiological noise reduction to 40% of precorrection magnitude, in agreement with published values (26), using an ASL excitation flip angle of 60 degrees, is plotted in Figure 5. The relative performance of the sequences is very similar for ASL (Fig. 5, top); for BOLD (Fig. 5, bottom), the double-excitation sequence performs slightly worse, although the reduction in BOLD CNR as compared to the double-echo sequence still only ranges between 3 and 6%.

Narrative Comprehension Task (Child Study)

The group BOLD and group ASL activation patterns (Fig. 6) both show activation in superior and middle temporal gyrus (BA 21/22) bilaterally, as shown previously for a BOLD-only version of this task in normal children (17,41). Z-scores (averaged over all voxels with significant group activation for both tasks) were 1.28 ± 0.73 for the BOLD data and 0.644 ± 0.67 for the ASL data. Normalizing to unity mean yielded an intersubject variance of 1.08 for the ASL data and 0.33 for the BOLD data. This difference is significant ($F(34,34) = 3.32, P < 0.001$). To remove the effect the age range and sex distribution of our subject population might have on the BOLD and ASL signals, we regressed out age and sex and recomputed the




Figure 3. Diagram of BOLD CNR (top) and CNR per unit time (bottom) as a function of slice number for the proposed double-excitation acquisition, a double-echo acquisition, and single-excitation single-echo acquisitions at varying TEs.

variances. In this case the intersubject variance was 0.99 for the ASL data and 0.29 for the BOLD data; this difference is also significant ($F(32,32) = 3.36$; $P < 0.001$). These results indicate significantly less intergroup variability and thus greater sensitivity for within-group analyses in the BOLD data.

Estimation of Physiological Noise Parameter at $TE = 11$ ms

Results (Table 2) show our results of $\lambda = .00596 \pm 0.0023$ in good agreement with our assumed theoretical value of $\lambda = 0.0059$. The physiological noise constant was higher ($\lambda = .00617 \pm 0.0023$) during the control acquisitions than during the labeled acquisitions ($\lambda = 0.00579 \pm 0.0023$) and this result was statistically significant ($P < 0.01$; paired t-test). As would be expected, the image SNR was higher during the control acquisition ($P < 0.001$; paired t-test), while the temporal SNR was lower ($P < 0.001$; paired t-test), as there is more signal during the control acquisitions from blood, a major contributor to physiological noise.

Empirical Comparison of Double-Excitation Versus Double-Echo Method (Adult Study)

Qualitatively, results look quite similar for the narrative comprehension task for the two techniques for BOLD (Fig. 7, top); however, there was much less activation detected with ASL using the double-echo technique (Fig. 7, bottom). Average T-scores for each

technique along with the CNR loss for the worse technique are listed in Table 3. A statistically significant difference was seen with ASL (paired t-test; $T = 4.82$; $P < 0.003$). For BOLD, the CNR loss is very insignificant at one percent or less, as expected from the simulation results. The ASL CNR loss for the double-echo technique (at around 40%) is somewhat larger than expected from the simulation (for slices acquired midway through the acquisition). However, in practice the CNR loss would be expected to be larger because the actual time needed for slice acquisition for the double-echo technique was greater than the value of $TE + 7$ ms assumed in the simulations. Complete agreement between experimental results and the simulations is also not expected due to differences in parameters such as tissue transit time, arterial transit time, and physiological noise constant.

DISCUSSION

A technique that simultaneously acquires ASL and BOLD with full-brain coverage using standard RF hardware has immediate applications for studies of brain development in children as well as for studying the developing child connectome. Future work will use this method to map growth trajectories of specific cognitive functions, with the goal to understand the relationship between developmental changes in perfusion and changes in neuronal activity associated with neurocognitive development. It is clearly necessary to image the whole brain to be able to understand the




Figure 4. Diagram of ASL CNR (top) and BOLD CNR (bottom) for the double-excitation acquisition as a function of slice number, using different excitation flip angles for the ASL acquisitions (excitation flip angle for the BOLD acquisitions remains at 90 degrees).

dynamics within and between brain networks and their impact on the developing brain, as is being investigated in the adult “Human Connectome” project, and to investigate the complex relationship between blood flow dynamics and neuronal function. In addition, this technique may also be useful in disease conditions where cerebral blood flow is atypical, such as stroke or cardiac disease.

The simulation results clearly show that the proposed double-excitation technique results in improved ASL functional contrast in whole-brain simultaneous BOLD/ASL acquisitions compared with previously described methods such as a double-echo method. By the time the 24th slice is acquired using a double-echo method, significant longitudinal recovery has occurred, resulting in only approximately 66% of the CNR remaining for ASL as compared to the double-excitation method. While, theoretically, the CNR available for the BOLD acquisition will be reduced as the effective TR is decreased, our simulations show that this cost is not great (only around 7%) and can be reduced even further (to only 1%) by using reduced excitation flip angles for the ASL acquisitions, with minimal cost in terms of ASL CNR.

In practice, our method may even provide slightly greater improvements in ASL CNR and smaller reductions in BOLD CNR compared with the double-echo

sequence. Our simulations did not take into account that for the double-echo sequence to acquire all the necessary data in the available time, it might be necessary to either increase acquisition bandwidth, or resort to partial k -space techniques, either of which will reduce the actual CNR. In fact, in actual practice we were not able to acquire slices as fast for the double-echo technique as was assumed in the simulations, for the same acquisition parameters as used for the double-excitation technique.

A different potential approach for simultaneous ASL/BOLD whole-brain imaging is to use the single-excitation single-echo technique. However, our results show this strategy to be sub-optimal, as using a single TE necessitates a compromise between BOLD CNR and ASL CNR. Physiological noise effects significantly reduce ASL CNR if a longer TE is used because the physiological noise constant λ increases with TE. This technique does benefit from a shorter minimum TR, and similar ASL CNR per unit time is available using TE = 15 ms compared with the double-excitation approach (Fig. 3, bottom); however, using this short TE results in a large reduction in BOLD CNR per unit time. Moreover, there may be practical limitations to this approach in pediatric populations and when pCASL labeling is accomplished using a body coil. The minimum TRs for the single-excitation single-




Figure 5. Diagram of ASL CNR (top) and BOLD CNR (bottom) as a function of slice number for the double-excitation acquisition, a double-echo acquisition, and single-excitation single-echo acquisitions at varying TEs, assuming that techniques for physiological noise reduction are successful in reducing the physiological noise to 40% of its precorrection magnitude.

echo sequences with TE = 15 ms and TE = 20 ms are under 3 s, and SAR limitations may render this short of a TR (with over half the time in each TR used for labeling) unfeasible in practice.

Our simulations also demonstrate the benefit of using reduced excitation flip angles for the ASL acquisitions. Our approach takes advantage of the fact that the physiological noise dominates the thermal noise for EPI acquisitions at these voxel sizes and TEs. This fact has also been used to

show that minimal reduction in CNR for BOLD fMRI occurs when excitation flip angles much less than the Ernst angle are used (19). Imaging at a lower flip angle may provide advantages such as enhanced tissue contrast and reduced inflow effects. Similarly, our approach reduces the excitation flip angles for the ASL acquisition, with minimal cost in terms of ASL CNR. This provides more longitudinal magnetization available for the BOLD scans.




Figure 6. Group functional activation (Left: ASL activation; Right: BOLD activation) from a cohort of 35 normal children ages 5–18 performing a narrative comprehension task using the double-excitation acquisition technique, demonstrating greater intergroup sensitivity in BOLD compared with ASL. All regions are significant at FWE corrected $P < 0.01$. Images in radiologic orientation (Slice locations: $Z = -47$ mm to $Z = +48$ mm, MNI coordinate space).

Table 2

Image SNR, Temporal SNR, and Physiological Noise Parameter λ for Each of 28 Children With Usable Data and the Group Average, Calculated From Baseline CBF Data

Subject	Combined			Control acquisitions only			Tagged acquisitions only		
	Temporal SNR	Image SNR	λ	Temporal SNR	Image SNR	λ	Temporal SNR	Image SNR	λ
1	176.330	824.260	0.00553988	169.592	828.933	0.00577179	183.069	819.587	0.00532440
2	159.696	677.948	0.00608569	158.747	679.955	0.00612523	160.644	675.941	0.00604658
3	150.552	531.756	0.00637045	164.965	534.298	0.00576573	136.139	529.214	0.00709823
4	113.984	527.537	0.00856595	107.588	529.982	0.00910120	120.379	525.092	0.00808583
5	155.234	705.653	0.00628410	152.620	708.801	0.00639853	157.847	702.504	0.00617324
6	108.237	461.438	0.00898126	97.1675	463.295	0.0100626	119.306	459.580	0.00809449
7	232.070	684.594	0.00405390	216.686	687.609	0.00437983	247.455	681.579	0.00376540
8	87.9514	534.047	0.0112147	90.9384	536.777	0.0108375	84.9645	531.316	0.0116182
9	134.017	423.561	0.00707838	129.461	424.935	0.00735712	138.573	422.187	0.00681661
10	219.887	550.070	0.00416863	211.533	552.004	0.00436652	228.241	548.135	0.00398344
11	202.709	657.368	0.00469279	185.251	660.194	0.00518120	220.166	654.542	0.00427737
12	94.2682	586.468	0.0104701	95.1375	589.834	0.0103735	93.3989	583.101	0.0105685
13	131.727	447.202	0.00725466	123.762	448.146	0.00776580	139.692	446.258	0.00679884
14	225.451	423.510	0.00375484	222.621	425.943	0.00382959	228.281	421.077	0.00368094
15	132.646	544.530	0.00731178	136.778	546.554	0.00707847	128.513	542.506	0.00755981
16	135.673	421.654	0.00697869	132.729	423.306	0.00715423	138.617	420.001	0.00680988
17	199.746	368.273	0.00420599	191.200	369.889	0.00447720	208.293	366.657	0.00395103
18	182.954	600.827	0.00520630	174.472	603.259	0.00548663	191.435	598.394	0.00494918
19	189.182	646.624	0.00505461	183.336	648.456	0.00523194	195.029	644.793	0.00488726
20	95.3438	773.090	0.0104083	88.5095	773.933	0.0112241	102.178	772.246	0.00970078
21	188.208	501.113	0.00492429	178.348	503.312	0.00524320	198.068	498.914	0.00463387
22	203.448	421.969	0.00430624	198.167	422.858	0.00445782	208.729	421.080	0.00416087
23	266.063	695.047	0.00347223	256.693	697.469	0.00362227	275.433	692.625	0.00333123
24	225.196	459.657	0.00387115	217.565	461.018	0.00405229	232.826	458.296	0.00369951
25	205.452	417.969	0.00423870	194.440	419.892	0.00455832	216.464	416.047	0.00394519
26	162.629	538.454	0.00586179	158.339	538.519	0.00603641	166.920	538.389	0.00569569
27	264.014	581.416	0.00337465	254.355	583.349	0.00353811	273.674	579.483	0.00322081
28	278.714	644.642	0.00323522	272.573	646.435	0.00332665	284.856	642.849	0.00314708
Average	175.764	558.953	0.00596304	170.128	561.034	0.00617156	181.400	556.871	0.00578658

*Values are averaged over control and tagged acquisitions (left), averaged over control acquisitions (center), and averaged over tagged acquisitions (right).

We note, however, that the effect of changing the effective TR is not identical to that of changing the flip angle. Unlike fully-relaxed spins sampled with a reduced flip angle, the proportion of signal coming from the blood will change if the effective TR is changed, because the T1 of the blood (which is the major contributor to the physiological noise) is different from the T1 of the surrounding gray matter. This will result in a nonlinear relationship between signal magnitude and physiological noise. However, the T1 of both arterial (20) and of venous blood (42), approximately 1600 ms, is slightly longer than the T1 of gray matter at 3T (43), approximately 1300 ms. This effect will make the amount of physiological noise *lower* than what would be predicted by the linear relationship used in our simulations. Thus, our simulations represent a lower bound of the actual BOLD CNR from the double-excitation sequence.

Our empirical estimation of the physiological noise parameter at TE = 11 ms of $\lambda = 0.0062$ during the control acquisitions and $\lambda = 0.0058$ during the labeled acquisitions yields good agreement with the results of (25), supporting our choice to use these published literature values in our CNR simulations. Our estimation is not intended to be a rigorous calculation of this parameter, as the estimation of the

signal noise using SENSE reconstruction has some bias associated with it. Our results, however, show that the literature values of physiological noise constants, typically obtained with single-channel coils or using sum-of-squares reconstruction with multichannel coils (19,25), are applicable to parallel imaging sequences, at least at larger voxel sizes, supporting the previous results of (34).

Results from the empirical comparison of the double-echo versus the double-excitation sequence (the adult study) agree nicely with our simulations. The results clearly show the loss of ASL signal using the double-echo technique, due to the time delay in acquiring those slices. Also in accordance with the simulations, minimal BOLD CNR loss results from using the double-excitation technique. Greater intra-subject BOLD CNR ($\sim 2-3 \times$) compared with ASL CNR, in agreement with previous studies (5,6), was also shown. This observation is subject to the caveat that BOLD CNR is not directly comparable to ASL CNR: the BOLD signal is highly autocorrelated due to the 1/f nature of the noise while the ASL control-label pairs approach statistical independence (5). On the other hand, with our implementation of the double excitation method, there are half as many control-label pairs available to compute functional ASL




Figure 7. Results (fixed-effects analysis) from a cohort of 7 normal adults performing a narrative comprehension task, comparing the double-echo acquisition technique to the double-excitation acquisition technique, showing minimal differences in BOLD sensitivity, but large differences in ASL sensitivity. Images in radiologic orientation (Slice locations: $Z = -4$ mm to $Z = +18$ mm, MNI coordinate space).

activation compared with total number of acquisitions used to compute functional BOLD activation. In practice, results will depend on the baseline CBF and CBF change from baseline for ASL contrast, and a range of additional factors including $CMRO_2$ and hematocrit for BOLD contrast, which will vary across populations and cortical regions.

Contrary to what has been suggested in a few previous reports (5,6), however, we found significantly greater sensitivity for within-group BOLD analyses compared with ASL in the child study. This discrepancy is likely related to methodology, sample size, and possibly activation paradigm. Aguirre et al (5) found greater group activation with ASL with a sample size of $N = 10$ in the visual cortex using a visual stimulation paradigm; however, a random-effects statistical test was not used, limiting generalizability. Moreover, this study used a sub-optimal BOLD scan protocol ($TE = 22$ ms at 1.5T) resulting in a BOLD CNR of approximately half of its optimal value. One study (6), using an optimized BOLD protocol, reported greater group activation with ASL, with a sample size of $N = 6$

adults in the motor cortex using a finger tapping paradigm, but only for task paradigms with greater than 4 min alternating on-off task periods. Our study involves a larger sample size of children ($N = 35$) and involves a different type of task (narrative processing) and activated region (superior temporal gyrus). Whether our results will generalize to different populations, tasks, and cortical regions is a subject for future research. Nevertheless, these results demonstrate the importance of optimizing combined ASL/BOLD acquisitions for ASL contrast.

A possible drawback of the proposed technique is that it is likely to increase inflow effects (44) for the BOLD acquisition in the first slices, due to the decreased effective TR. Although inflow effects have the desirable property of increasing overall contrast in BOLD scans, they have the undesirable property that they are located primarily in large arteries and draining veins, away from the sites of neuronal activation. This effect, however, is likely not of major concern for the voxel sizes typically used on whole-brain ASL/BOLD ($\sim 4 \times 4 \times 4$ mm). Moreover, inflow effects for

Table 3
Empirical Comparison of Double-Excitation Acquisition Technique to Double-Echo Technique^a

Contrast	T-score (mean \pm SEM), double-excitation	T-score (mean \pm SEM), double-echo	<i>P</i>	CNR loss, % (mean \pm SEM)
ASL	2.75 \pm 0.38	1.66 \pm 0.43	.003	42.51 \pm 8.25
BOLD	4.44 \pm 0.86	4.57 \pm 0.82	n.s.	0.54 \pm 16.03

^aAverage T-scores and CNR loss for the worse-performing technique for ASL and BOLD contrasts for the narrative comprehension task in a cohort of seven normal adults, across the ROIs shown in Figure 7. *P* values obtained using paired t-test.

n.s. = not significant.

gradient-echo EPI sequences appear to primarily affect the shape and latency of the hemodynamic response function, as opposed to activation magnitude (45). Nevertheless, if inflow effects are a concern, they can be significantly reduced by lowering the flip angle of the excitation pulses for the BOLD acquisitions, as previously proposed for BOLD-only acquisitions (19), and is a subject for future research.

Additionally, while our ASL acquisitions did not use crusher gradients to null signal from the arteries and confine it to arterioles and capillaries, such gradients may be deployed if confinement of contrast to the brain parenchyma is desired, with the tradeoff of reduced CNR. Note that ASL contrast is primarily present on the arterial side and in the capillaries, as contrast in the veins is severely reduced due to the short T1 of blood. An alternative strategy to localize contrast to the brain parenchyma is to increase the postinversion delay such that all spins will have diffused into the tissue compartment; this also comes at the cost of reduced CNR and also of increased TR.

We note that our implementation is not specific to 2D multislice acquisitions, and a possible alternative is a 3D technique such as single-shot 3D GRASE (46) or a stack of spirals 3D acquisition (47). As originally proposed, the 3D GRASE acquisition uses a repeated short ($> \sim 20$ ms) 2D EPI readout for each k_z , with refocusing RF pulses between each readout. Like our proposed double-excitation sequence, acquisitions are kept as close to the tagging pulse as possible for optimal ASL contrast. The 3D GRASE acquisition is not optimal for BOLD because the contrast is predominantly T2, not T2*, weighted. However, a T2*-weighted 3D acquisition such as 3D PRESTO (48,49) could follow the GRASE acquisition for optimized BOLD contrast. Similar to what was done in the multislice 2D case, BOLD CNR can be improved by reduction of the ASL excitation flip angle (to preserve longitudinal magnetization for the BOLD scan). Readout acquisition time can be shortened by means of parallel imaging and/or partial k -space techniques (possibly in both phase encoding directions). Comparison of 3D to multislice acquisition will likely depend on available hardware and a detailed comparison of the two techniques is beyond the scope of this study.

In conclusion, the ASL/BOLD functional imaging method proposed here will allow nearly simultaneous acquisition of perfusion-weighted and BOLD-weighted functional maps from the same task. We are currently undertaking a large scale study using this method in children ranging in age from birth to 18 years with the goal in mind of mapping developmental trajectories in brain activity and connectivity in this age group. However, we would like to point out that the proposed technique greatly facilitates a broad range of other applications. The quantitative nature of the technique can provide a stable marker of brain activity, unlike BOLD imaging alone, which is hampered by scanner drifts and variance of the BOLD response across subjects and scanners (5,6,50,51). This technique can thus be a very powerful for longitudinal functional MRI studies, multicenter studies or studies comparing subject populations. Such applications could include drug

studies, studies of cortical plasticity as a result of cognitive training (52,53), or examining brain function differences due to psychiatric disorders.

REFERENCES

- Schapiro MB, Schmithorst VJ, Wilke M, Byars AW, Strawsburg RH, Holland SK. BOLD fMRI signal increases with age in selected brain regions in children. *Neuroreport* 2004;15:2575–2578.
- Toga AW, Clark KA, Thompson PM, Shattuck DW, Van Horn JD. Mapping the human connectome. *Neurosurgery* 2012;71:1–5.
- Hoge RD. Calibrated fMRI. *Neuroimage* 2012;62:930–937.
- Hoge RD, Atkinson J, Gill B, Crelier GR, Marrett S, Pike GB. Investigation of BOLD signal dependence on cerebral blood flow and oxygen consumption: the deoxyhemoglobin dilution model. *Magn Reson Med* 1999;42:849–863.
- Aguirre GK, Detre JA, Zarahn E, Alsop DC. Experimental design and the relative sensitivity of BOLD and perfusion fMRI. *Neuroimage* 2002;15:488–500.
- Wang J, Aguirre GK, Kimberg DY, Roc AC, Li L, Detre JA. Arterial spin labeling perfusion fMRI with very low task frequency. *Magn Reson Med* 2003;49:796–802.
- Woolrich MW, Chiarelli P, Gállichan D, Perthen J, Liu TT. Bayesian inference of hemodynamic changes in functional arterial spin labeling data. *Magn Reson Med* 2006;56:891–906.
- Liu TT, Wong EC, Frank LR, Buxton RB. Analysis and design of perfusion-based event-related fMRI experiments. *Neuroimage* 2002;16:269–282.
- Glielmi CB, Xu Q, Craddock RC, Hu X. Simultaneous acquisition of gradient echo/spin echo BOLD and perfusion with a separate labeling coil. *Magn Reson Med* 2010;64:1827–1831.
- Kastrup A, Kruger G, Neumann-Haefelin T, Glover GH, Moseley ME. Changes of cerebral blood flow, oxygenation, and oxidative metabolism during graded motor activation. *Neuroimage* 2002;15:74–82.
- Stefanovic B, Warnking JM, Rylander KM, Pike GB. The effect of global cerebral vasodilation on focal activation hemodynamics. *Neuroimage* 2006;30:726–734.
- St Lawrence KS, Frank JA, Bandettini PA, Ye FQ. Noise reduction in multi-slice arterial spin tagging imaging. *Magn Reson Med* 2005;53:735–738.
- Ye FQ, Frank JA, Weinberger DR, McLaughlin AC. Noise reduction in 3D perfusion imaging by attenuating the static signal in arterial spin tagging (ASSIST). *Magn Reson Med* 2000;44:92–100.
- de Bie HM, Boersma M, Adriaanse S, et al. Resting-state networks in awake five- to eight-year old children. *Hum Brain Mapp* 2012;33:1189–1201.
- Karunanayaka P, Schmithorst VJ, Vannest J, Szaflarski JP, Plante E, Holland SK. A group independent component analysis of covert verb generation in children: a functional magnetic resonance imaging study. *Neuroimage* 2010;51:472–487.
- Rubia K. Functional brain imaging across development. *Eur Child Adolesc Psychiatry* 2012 [Epub ahead of print].
- Schmithorst VJ, Holland SK, Plante E. Cognitive modules utilized for narrative comprehension in children: a functional magnetic resonance imaging study. *Neuroimage* 2006;29:254–266.
- Schmithorst VJ, Holland SK, Plante E. Object identification and lexical/semantic access in children: a functional magnetic resonance imaging study of word-picture matching. *Hum Brain Mapp* 2007;28:1060–1074.
- Gonzalez-Castillo J, Roopchansingh V, Bandettini PA, Bodurka J. Physiological noise effects on the flip angle selection in BOLD fMRI. *Neuroimage* 2011;54:2764–2778.
- Wang J, Alsop DC, Li L, et al. Comparison of quantitative perfusion imaging using arterial spin labeling at 1.5 and 4.0 Tesla. *Magn Reson Med* 2002;48:242–254.
- Gonzalez-At JB, Alsop DC, Detre JA. Cerebral perfusion and arterial transit time changes during task activation determined with continuous arterial spin labeling. *Magn Reson Med* 2000;43:739–746.
- Lu H, Clingman C, Golay X, van Zijl PCM. Determining the longitudinal relaxation time (T1) of blood at 3.0 Tesla. *Magn Reson Med* 2004;52:679–682.
- Wang J, Alsop DC, Song HK, et al. Arterial transit time imaging with flow encoding arterial spin tagging (FEAST). *Magn Reson Med* 2003;50:599–607.

24. Yoshiura T, Hiwatashi A, Yamashita K, et al. Simultaneous measurement of arterial transit time, arterial blood volume, and cerebral blood flow using arterial spin-labeling in patients with Alzheimer disease. *AJNR Am J Neuroradiol* 2009;30:1388–1393.
25. Kruger G, Glover GH. Physiological noise in oxygenation-sensitive magnetic resonance imaging. *Magn Reson Med* 2001;46:631–637.
26. Glover GH, Li TQ, Ress D. Image-based method for retrospective correction of physiological motion effects in fMRI: RETROICOR. *Magn Reson Med* 2000;44:162–167.
27. Dinov I, Lozev K, Petrosyan P, et al. Neuroimaging study designs, computational analyses and data provenance using the LONI pipeline. *PLoS One* 2010;5:pii: e13070.
28. Dinov ID, Van Horn JD, Lozev KM, et al. Efficient, Distributed and Interactive Neuroimaging Data Analysis Using the LONI Pipeline. *Front Neuroinform* 2009;3:22.
29. Thevenaz P, Ruttimann UE, Unser M. A pyramid approach to subpixel registration based on intensity. *IEEE Trans Image Process* 1998;7:27–41.
30. Szaflarski JP, Schmithorst VJ, Altaye M, et al. A longitudinal functional magnetic resonance imaging study of language development in children 5 to 11 years old. *Ann Neurol* 2006;59:796–807.
31. Wilke M, Holland SK, Altaye M, Gaser C. Template-O-Matic: a toolbox for creating customized pediatric templates. *Neuroimage* 2008;41:903–913.
32. Wong EC, Buxton RB, Frank LR. Implementation of quantitative perfusion imaging techniques for functional brain mapping using pulsed arterial spin labeling. *NMR Biomed* 1997;10:237–249.
33. Liu TT, Wong EC. A signal processing model for arterial spin labeling functional MRI. *Neuroimage* 2005;24:207–215.
34. Triantafyllou C, Polimeni JR, Wald LL. Physiological noise and signal-to-noise ratio in fMRI with multi-channel array coils. *Neuroimage* 2011;55:597–606.
35. Gilbert G. Measurement of signal-to-noise ratios in sum-of-squares MR images. *J Magn Reson Imaging* 2007;26:1678.
36. Dietrich O, Raya JG, Reeder SB, Reiser MF, Schoenberg SO. Measurement of signal-to-noise ratios in MR images: influence of multichannel coils, parallel imaging, and reconstruction filters. *J Magn Reson Imaging* 2007;26:375–385.
37. Jenkinson M, Bannister P, Brady M, Smith S. Improved optimization for the robust and accurate linear registration and motion correction of brain images. *Neuroimage* 2002;17:825–841.
38. Mumford JA, Hernandez-Garcia L, Lee GR, Nichols TE. Estimation efficiency and statistical power in arterial spin labeling fMRI. *Neuroimage* 2006;33:103–114.
39. Woolrich MW, Ripley BD, Brady M, Smith SM. Temporal autocorrelation in univariate linear modeling of FMRI data. *Neuroimage* 2001;14:1370–1386.
40. Avants BB, Epstein CL, Grossman M, Gee JC. Symmetric diffeomorphic image registration with cross-correlation: evaluating automated labeling of elderly and neurodegenerative brain. *Med Image Anal* 2008;12:26–41.
41. Vannest JJ, Karunanayaka PR, Altaye M, et al. Comparison of fMRI data from passive listening and active-response story processing tasks in children. *J Magn Reson Imaging* 2009;29:971–976.
42. Wu WC, Jain V, Li C, et al. In vivo venous blood T1 measurement using inversion recovery true-FISP in children and adults. *Magn Reson Med* 2010;64:1140–1147.
43. Wansapura JP, Holland SK, Dunn RS, Ball WS Jr. NMR relaxation times in the human brain at 3.0 tesla. *J Magn Reson Imaging* 1999;9:531–538.
44. Gao JH, Liu HL. Inflow effects on functional MRI. *Neuroimage* 2012;62:1035–1039.
45. Liu HL, Wei PS, Wai YY, et al. Inflow effects on hemodynamic responses characterized by event-related fMRI using gradient-echo EPI sequences. *Med Phys* 2008;35:4300–4307.
46. Fernandez-Seara MA, Wang Z, Wang J, et al. Continuous arterial spin labeling perfusion measurements using single shot 3D GRASE at 3 T. *Magn Reson Med* 2005;54:1241–1247.
47. Nielsen JF, Hernandez-Garcia L. Functional perfusion imaging using pseudocontinuous arterial spin labeling with low-flip-angle segmented 3D spiral readouts. *Magn Reson Med* 2013;69:382–390.
48. Golay X, Pruessmann KP, Weiger M, et al. PRESTO-SENSE: an ultrafast whole-brain fMRI technique. *Magn Reson Med* 2000;43:779–786.
49. van Gelderen P, Duyn JH, Ramsey NF, Liu G, Moonen CT. The PRESTO technique for fMRI. *Neuroimage* 2012;62:676–681.
50. McGonigle DJ, Howseman AM, Athwal BS, Friston KJ, Frackowiak RS, Holmes AP. Variability in fMRI: an examination of inter-session differences. *Neuroimage* 2000;11(Pt 1):708–734.
51. Smith AM, Lewis BK, Ruttimann UE, et al. Investigation of low frequency drift in fMRI signal. *Neuroimage* 1999;9:526–533.
52. Buschkuehl M, Jaeggi SM, Jonides J. Neuronal effects following working memory training. *Dev Cogn Neurosci* 2012;2(Suppl 1):S167–S179.
53. Jaeggi SM, Buschkuehl M, Jonides J, Shah P. Short- and long-term benefits of cognitive training. *Proc Natl Acad Sci U S A* 2011;108:10081–10086.