

DOI: 10.1002/((please add manuscript number))

Article type: Communication

Ultra-Specific Isolation of Circulating Tumor Cells Enables Rare-Cell RNA Profiling

*Rhonda M. Jack[†], Meggie M. G. Grafton[†], Danika Rodrigues, Maria D. Giraldez, Catherine Griffith, Robert Cieslak, Mina Zeinali, Chandan Kumar-Sinha, Ebrahim Azizi, Max Wicha, Muneesh Tewari, Diane M. Simeone, and Sunitha Nagrath**

R. M. Jack, Dr. M. M. G. Grafton, C. Griffith, R. Cieslak, M. Zenali, Prof. S. Nagrath
Department of Chemical Engineering, 3074 H.H. Dow Building, 2300 Hayward St., Ann Arbor, MI
48109-2136, USA

E-mail: snagrath@umich.edu

R. M. Jack, Dr. M. M. G. Grafton, M. Zenali, Dr. M. Tewari, Prof. S. Nagrath

Biointerfaces Institute, North Campus Research Complex, 2800 Plymouth Road, Bldg. 90, Ann Arbor,
MI 48109-2800, USA

D. Rodrigues, Dr. M. Tewari
Department of Biomedical Engineering, 1107 Carl A. Gerstacker Building, 2200 Bonisteel, Blvd., Ann
Arbor, MI 48109, USA

Dr. M. Giraldez, Dr. M. Tewari
Department of Internal Medicine, Division of Hematology/Oncology and Division of Molecular
Medicine and Genetics, University of Michigan, 109 Zina Pitcher Place, 2061 BSRB - SPC 2200, Ann
Arbor, MI 48109, USA

M. Zeinali
Medical Faculty, University of Heidelberg, Theodor-Kutzer-Ufer 1-3 68167 Mannheim, Germany

Dr. C. Kumar-Sinha
Michigan Center for Translational Pathology, Department of Pathology, University of Michigan 5309
CCC 5940, 400 E. Medical Center Dr., Ann Arbor, MI 48109-0940

This is the author manuscript accepted for publication and has undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the [Version of Record](#). Please cite this article as [doi: 10.1002/advs.201600063](https://doi.org/10.1002/advs.201600063).

This article is protected by copyright. All rights reserved.

Dr.E. Azizi, Dr. M. Wicha, Dr. M. Tewari

Comprehensive Cancer Center, University of Michigan Health Systems, 1500 E. Medical Center Dr,
Ann Arbor, MI 48109, USA

Dr. M. Tewari

Center for Computational Medicine and Bioinformatics, University of Michigan, Room 2017, Palmer
Commons, 100 Washtenaw Avenue, Ann Arbor, MI 48109-2218

Dr. D. M. Simeone

Department of Surgery, University of Michigan Health Systems, 1500 E. Medical Center Dr, Ann
Arbor, MI 48109, USA

Keywords: cell sorting, immunomagnetism, microfluidics, microRNA, tumor cells

While substantial evidence exists to support the clinical potential of circulating tumor cells (CTCs) in elucidating metastasis, achieving high CTC isolation purities from blood remains a challenge. Herein we report an ultra-specific microfluidic CTC sorting approach that enables significant CTC enrichment in a continuous, high-throughput manner (24 mL hr^{-1}), through strategic combination of size based-inertial and magnetic CTC sorting. An unequalled purity rate, averaged at $82.5\% \pm 24.3$ was achieved across 14 pancreatic ductal adenocarcinoma (PDAC) patients, with $>13 \text{ CTCs mL}^{-1}$ observed in 100% of samples (median = 69 CTCs mL^{-1} , mean = $146 \pm 231 \text{ CTCs mL}^{-1}$). Both MicroRNA and mRNA profiling of CTCs revealed that RNAs previously reported as PDAC-associated, were among those most abundantly expressed. In summary, we describe a microfluidic platform that enhances CTC purity and permits detailed molecular characterization of isolated CTCs.

Not surprisingly CTC biology has sparked burgeoning interest within recent years because of the postulated role of these rare cells in metastasis. CTCs initiate from the primary tumor and migrate in the blood then occupy a secondary site to subsequently metastasize.¹ They inherently carry information about the tumor that could be useful for understanding metastasis and developing novel therapeutics to more effectively treat these potentially harmful cells. Since CTCs can be isolated from a simple blood draw, they can be routinely

used to yield insight into tumor pathology and treatment response without invasive procedures.²

However CTC detection and isolation from whole blood is quite challenging. Firstly, CTCs are rare; there may be only 1 CTC in 7.5mL of blood containing billions of blood cells. Secondly, there is limited knowledge of specific CTC markers, due in part to CTC heterogeneity such that they vary phenotypically and genotypically among patients.³ Considering CTC rarity and heterogeneity, engineering a device to isolate them from blood presents an interesting, important challenge.

Two chief approaches used for CTC collection are immunoaffinity capture and label free, size-based filtering.⁴ Immune-based methods commonly rely on expression of epithelial cell adhesion molecule (EpCAM), a surface protein, to isolate CTCs and typically require low flow rates (1-3mL hr⁻¹) to ensure that CTCs have sufficient time to be immobilized by functionalized surfaces.⁵⁻⁶ Issues concerning cell release of the immobilized cells also arise.⁷ In contrast label-free isolation does not rely on specific biomarkers but instead on inherent CTC properties such as size, deformability, or dielectric susceptibility. Size-based sorting usually offers higher throughputs since bulk sorting is used.

Moreover, since the discovery of microfluidic inertial size-based sorting, it was reported that enhanced sorting is achieved by incorporating curvature into microfluidic channels. Curvature introduces Dean Forces that augment the inertial forces acting on particles under flow and can generally be performed with high flow rates ranging from 500-2000 $\mu\text{L min}^{-1}$, while maintaining typical microfluidic dimensions (see Supplementary Information (SI) Section 2).⁸⁻¹¹ Whereas label-free methods are rapid, specificity is lacking and operation is

limited by overall high blood cell counts. The recently reported CTC iChip uses hydrodynamic, inertial and magnetic sorting to isolate CTCs from whole blood.¹² A throughput of 8mL hr⁻¹ using positive immunomagnetic selection, with an average of 1500 contaminating white blood cells (WBCs) from 1 mL whole blood were achieved. Purity was affected by non-specific WBC labeling with magnetic beads.

We report the development of an ultra-specific microfluidic device that combines inertial microfluidics and immuno-magnetism in a uniquely designed workflow that minimizes non-specific cell labeling for efficient isolation of CTCs from whole blood with ultra-high CTC purity rates and high throughput. We hypothesized that beginning the CTC isolation process by rapid inertial pre-sorting would greatly reduce the number of unwanted cells from whole blood so that micron-scale, on-chip passive mixing and brief on-chip incubation would facilitate specific, extensive CTC magnetic bead labeling. Then, ultra-specific sorting of the pre-enriched magnetized CTCs could be achieved via magnetic sorting. Specifically, the microfluidic design would allow for whole blood to be processed rapidly yet meticulously on a single chip, requiring no blood preparation or long incubation times. This would minimize the extent of non-specific binding to promote CTC isolation purity. Unlike most other platforms, our approach negates bulk CTC labeling and instead uses on-chip CTC-bead labeling which improves the labeling chances of target cell magnetization and successful isolation from blood samples (see **table S1**).

The functionality of each component is illustrated in **Figure 1**, showing cell distribution throughout the device. The first step of inertial sorting involves flowing whole blood into one inlet and phosphate buffered saline (PBS) into the other inlet of the spiral module. The use of

two differently sized syringes biases the two flows such that whole blood flows at approximately $400 \mu\text{L min}^{-1}$, while buffer flows at $1200 \mu\text{L min}^{-1}$. By biasing the flow the higher buffer flow rate readily stabilizes flow prior to the blood sample introduction so that no blood sample is wasted initially due to the flow stabilization process. Concurrently the blood sample is diluted so that the effect of waste cells on cancer cell flow dynamics, as well as on purity and recoveries is effectively reduced.

Next, the CTC-enriched stream which leaves the inertial sorter at the innermost outlet, enters the passive mixer where it mixes with EpCAM coated magnetic beads infused at $100 \mu\text{L min}^{-1}$. After mixing, 5 min on-chip incubation is carried out in four reservoirs to promote extensive magnetic bead-CTC labeling via antibody-antigen interactions.

To achieve immediate CTC magnetic sorting after incubation, the reservoir inlet is gently and gradually pressurized with an air-filled syringe so that the mixture can flow out of the reservoirs in a controlled manner. The mixture drains directly into the magnetic sorter between two buffer streams and permanent magnets lined along the magnetic sorter deflect the magnetically labeled CTCs away from unlabeled blood cells to the CTC collection outlet.

The final collection is a highly enriched suspension of viable CTCs.

The inertial sorter separates larger particles ($15 -20\mu\text{m}$) from smaller particles ($7-10\mu\text{m}$).

Figure 2A shows the fabricated polydimethylsiloxane (PDMS) device, figure 2B illustrates the arrangement of cancer cells (green) versus WBCs (blue) at the outlet of the sorter and figure 2C demonstrates how the polystyrene particles ($20 \mu\text{m}$ (green), $15 \mu\text{m}$ (blue), $7 \mu\text{m}$ (orange)) separate into distinct, focused streams at the outlets of the spiral channel. Similarly in figure 2D PANC-1 cancer cells (green streak) focus to the innermost outlet and WBCs

(blue streak) focus to the second outlet. Figure 2E shows how the larger PANC-1 cells (indicated by red arrows) spiked into diluted blood focus to the innermost outlet amidst the other blood cells.

To quantitatively evaluate inertial sorting PANC-1 cells and WBCs were spiked into PBS and processed through the module. The effluents from each outlet were collected and analyzed for cell counts. Notably, figure 2F shows $90.95\% \pm 3.61$ of cancer cells were collected in the desired outlet. Whole blood was routinely diluted by x10 for blood-spiked experiments to facilitate conditions amenable to microscope visualizing. Figure 2G shows cell distribution in the four outlets after repeating the experiments in diluted blood. Results consistent with PBS-spiked experiments were achieved such that $91.99\% \pm 4.94$ of cancer cells were enriched in the innermost spiral outlet. The spiral design was optimized as a stand-alone device and subsequently, the three waste outlets were re-designed to allow for resistance-matching once it was connected within the fully integrated system (see **figure S1** for and **table S2**).

The passive mixer facilitates mixing of EpCAM magnetic beads with the CTC-rich effluent from the spiral sorter so that magnetic beads bind to cancer cells through antibody-antigen complexes. Mixing occurs after inertial whole blood pre-sorting to increase the chances of bead-cancer cell interactions, which economizes on the quantity of required antibodies. Additionally, mixing occurs at length-scales comparable to cancer cell sizes thus promoting efficient bead-CTC labelling compared to bulk labelling. Because both viscous and inertial forces are appreciable in this range of Re , “secondary flows” arise in both axial and radial directions thus increasing the interfacial area over which diffusion and mixing occur (see SI

Section 3).¹³ Furthermore, the short mixing time required decreases the likelihood of nonspecific binding with non-target cells.

Using flow rates of 10-100 $\mu\text{L min}^{-1}$, mixers were first evaluated by testing the distance along various designs required for two distinct streams of yellow- and blue- dyed PBS to become one well-mixed green stream (see SI Figure S1). The highly compact selected mixer design enabled fluids to completely mix by the third channel segment, after $\sim 4\text{cm}$ footprint. Next cancer cell labeling with EpCAM beads was examined through static experiments (PANC-1 cells and EpCAM beads in an Eppendorf tube) and on-chip experiments (on-chip passive mixing of cells and beads subsequently flowed into reservoirs). Calculations demonstrate that for a 20 μm cell to experience a sufficient magnetic force to pull it in the direction perpendicular to flow and therefore sort it, less than 1/3 bead coverage is required (see SI Section 4). For four incubation time points (5, 10, 15 and 30 min), labeling was evaluated by immediately imaging the mixed effluent to quantify beads attached to cells. Cells were characterized as having no coverage, $<1/3$, $1/3-2/3$ and $> 2/3$ coverage (see SI Figure S3). In static conditions 15 min incubation allowed at least 1/3 coverage on only 85% of cells (SI Figure S6). Conversely when cells diluted in blood and beads were mixed on-chip at $100\mu\text{L min}^{-1}$ and allowed 5 min incubation in the reservoirs, $>92\%$ of the cells demonstrated over 1/3 coverage. Thus, 5 minutes of incubation was determined to be optimal for on-chip labeling. Not only do the reservoirs facilitate on-chip incubation for extensive cancer cell bead coverage, yet it enables a seamless switch of flow rates from $1200 \mu\text{L min}^{-1}$ to $50 \mu\text{L min}^{-1}$, which in turn facilitates sensitive, thorough magnetic sorting of magnetized cancer cells.

The magnetic sorter has three inlets consisting of 2 outer buffer streams and 1 middle sample stream that flow in a 250 μm -wide channel and allows for further continuous on-chip cell sorting.¹⁴ Confining the sample stream in this manner ultimately provides efficient magnetized CTC sorting from contaminating cells by ensuring that only magnetized cells are attracted into the collection outlet. The sorter outlet is split into 2 channels, with a smaller collection (35 μm) and a larger waste (65 μm) channel. Permanent magnets (500mT, K&J Magnetics) aligned end-to-end, are positioned parallel to the main channel with a 1 mm PDMS separation gap between the channel and the magnets. The sample continuously flows through the sorter at 50 $\mu\text{L min}^{-1}$ so that cells with as few as 1/3 surface bead coverage are deflected into the collection channel. Waste cells continue uninterrupted to the waste outlet.

In determining the appropriate magnetic sorting flow rates, PANC-1 cells pre-labeled with magnetic beads were flowed through the sorter at test rates of 25-100 $\mu\text{L min}^{-1}$, with top (farther from magnet) buffer streams of 15-50 $\mu\text{L min}^{-1}$ and bottom (closer to magnet) buffer streams of 25-75 $\mu\text{L min}^{-1}$. Biasing the flow rates was readily achieved by using two different size syringes at a fixed syringe drive setting so that the top and bottom streams could simultaneously flow at two different flow rates. Notably, biasing the flow in this manner significantly improved magnetic sorting by allowing magnetized cells to flow slower to maximize magnetic trapping and to enhance separation distance of waste cells from the collection channel (see Video S1). Using 75 $\mu\text{L min}^{-1}$ - bottom, 25 $\mu\text{L min}^{-1}$ -top buffer and 50 $\mu\text{L min}^{-1}$ flow rates, results show that for blood spiked tests, WBC contamination was 2.28%

± 0.68 and recovery efficiency was $94.60\% \pm 1.48$. Of note, the few PANC-1 cells in the waste were poorly labeled and so not expected to be magnetically sorted (Figure S6).

To establish the proof of concept of the device, 1 mL whole blood was routinely processed. PANC-1 cells spiked at 100 and 1000 cells mL^{-1} were used to evaluate device performance with cancer cells. Results demonstrate that efficiency of PANC-1 collection was nearly 90% (Figure 3C). For both cases, contamination ranged from 82-801 WBCs mL^{-1} (Figure 3D), resulting in purities of up to 75%. The MTT assay was used to assess isolated PANC-1 cell viability which revealed that cells processed through the integrated system had viability rates comparable to that of the unprocessed control cells thus supporting the feasibility of culturing the collected cells (Figure S8). Also, in consideration of developing a point-of-care diagnostic device, the antibody-labeled beads were tested and demonstrated stability in cancer cell labeling even after 100 days (see SI Section 7). The unprecedented purity rates achieved by the device supersedes some of the other reported technologies including GEDI (68%), herringbone (14%), iChip (1500 WBCs mL^{-1}), and the CTC chip (9%)^{5, 12, 15-16} and facilitates miRNA profiling.

Fourteen metastatic or locally advanced PDAC patient samples and 4 healthy controls were processed through the device (Figure 4A). Up to 1.4 mL of whole blood from each patient was processed and CTCs were identified via fluorescent immunostaining for CK+/DAPI+/CD45-. Additional confirmation of an epithelial cell type was garnered by brightfield images showing bound EpCAM magnetic beads on the cells. WBCs were identified as CK-/DAPI+/CD45+. Figure 4D and 5E are representative images of fluorescently labeled CTCs and WBCs respectively. Samples were enumerated and CTC

purity percentages are as reported in Figure 5C. The corresponding patient clinical data is shown in Table S1. For all 14 patients, CTCs were isolated with counts ranging from 14-938 CTCs mL⁻¹ which was higher than the average 3 CTCs mL⁻¹ for healthy controls. Notably, CTCs have been recovered from patient blood with significant depletion, on the order of 10⁶ WBCs with average CTC purity being 82.5% ±23.5.

Although magnetic sorting is the rate limiting process, occurring at 50 μL min⁻¹, it can easily be multiplexed. Using parallel sorters (n=8) on a single chip, the entire process can be maintained at 400 μL min⁻¹, yielding an overall processing rate of greater than 20mL hr⁻¹ and still yield unprecedented purities of greater than 80%.

MiRNAs are important regulators of cancer cell properties however they have been in general, understudied in the context of CTCs. Currently, the relatively low purity rates proffered by most CTC isolation technologies limit specificity in genomic characterizing. The high CTC purities of our device enable CTC-specific miRNA profiling. Accordingly we profiled 372 miRNAs in CTC enriched samples from 2 locally advanced PDAC patients alongside a healthy control individual.

Of the 372 miRNAs profiled, 106 miRs demonstrated abundance in expression in both samples compared to the healthy control. 81 miRs had low expression and were categorized as not reliably detected. Interestingly, several miRs previously reported as associated with PDAC including, miR-221-3p, 210, -23a-3p, -143-3p and -21-5p were among those highly expressed in the CTC samples (Figure 4 c).¹⁷⁻²⁰ Additionally, based on pathway analysis, the top network functions enriched in the profiles were for cancer and organismal injury and abnormalities. Pathways involving Smad2/3 and MAP2K1/2 were implicated in PDAC CTC

biology (see SI Figure S9). Moreover, for the first time we report that PDAC CTC miRNA profiling revealed a trend towards high expression of miRs 17-5p, 19b-3p, 320b and let7a-5p. Also, miR-96-3p, 216b, 155-5p, 212-3p and -31-5p which demonstrated decreased expression have been previously reported to exhibit down regulation in pancreatic cancer.¹⁷⁻¹⁸

To further demonstrate the potential of the integrated device in allowing sensitive molecular interrogation of CTC samples subsequent to CTC isolation, mRNA profiling was carried out on eight additional patient samples. Of these, three were categorized as borderline and two as locally advanced while three were metastatic. When metastatic profiles were grouped against other patients, as seen in **figure 5A**, all borderline profiles clustered closely while two of three metastatic cases clustered apart from all other profiles. Moreover, seven mRNAs Zeb1, KLF4, cMYC, BRCA1, HER2, CD3D, and HPRT1 levels (indicated by the green box in figure 5A) were able to distinguish among the two groups. SPARC was also higher-expressed in the metastatic group (LogFC=4.7, p=0.037). Specifically, Zeb1 (LogFC=11.6, p=0.037) has shown to be an independent predictor of mortality in pancreatic cancer while the cause of overexpression of cMYC (LogFC=9.9, p=0.037) in pancreatic cancers has been previously studied.²¹⁻²² The role of KLF4 (LogFC=11.6, p=0.037) as both as a tumor suppressor and an oncogene in cancer progression was previously reported.²³ Interestingly, BRCA1 (LogFC=10.9, p=0.037) was reported to be down regulated in cases of pancreatitis and sporadic PDAC.²⁴ Additionally the results of HER2 amplification (LogFC=10.1, p=0.037) among PDAC patients showed potential for therapeutic endeavors.²⁵ Aberrantly high expression of SPARC (LogFC=4.7, p=0.037) in pancreatic CTCs was also observed in both mouse and human samples.²⁶ Taken together, the corroboration of reported studies by the

results of CTC molecular profiling enriched by the described device show promise in its use to ascertain information that may be specific to CTCs.

In summary, the integrated microfluidic device described is capable of generating significantly enriched, viable CTCs which enable CTC miRNA and mRNA analysis. The device uses inertial sorting to perform rapid pre-sorting of the original sample which significantly reduces the amount of sample, thus reducing the overall time required for processing. On-chip magnetic labeling allows for very sensitive CTC-bead tagging in the absence of the majority of blood cells and reduces non-specific binding. By multiplexing the magnetic sorter module, the device can process up to 10 mL of blood in 1 hour and is a closed device requiring limited user involvement. Ultra-pure CTC-specific miRNA profiling revealed abundance in expression of miRs 17-5p, 19b-3p, 320b and let7a-5p along with mRNAs Zeb1, KLF4, cMYC, BRCA1, HER2, CD3D, and HPRT1 in PDAC derived CTCs. The platform presented here should enable more studies of CTC RNA expression which can in turn, help to identify their putative roles in cancer metastasis.

Supporting Information


Supporting Information is available from the Wiley Online Library or from the author.

[†]Authors contributed equally to this work.

Acknowledgements

This article is protected by copyright. All rights reserved.

This research was funded by DOD-CDMRP-PCRP Concept Award W81XWH-12-PCRP-CDA to S. Nagrath, National Institutes of Health (NIH) Transformative R01 Grant R01DK085714 to M. Tewari, and Rio Hortega and Martin Escudero fellowships to M. Giraldez. This research was also funded by the Lustgarten Foundation, as well as a generous donation from the Sky Foundation (S. Nagrath and D. Simeone). We thank H. Cameron for patient sample retrieval, and H. J. Yoon for microfabrication training. We thank D. Jue for assistance with cell staining optimization. We also thank the Lurie Nanofabrication Facility, a member of the National Nanotechnology Infrastructure Network (NNIN) funded by the National Science Foundation for use of equipment to fabricate devices.



This article is protected by copyright. All rights reserved.

Figure 1. Design of ultra-Specific CTC isolation platform (A) Schematic showing inertial sorting, passive mixing with incubation and magnetic sorting (B) 3D illustration of PDMS device (C) Reservoirs allow for brief incubation and flowrate reduction to enable sensitive magnetic sorting.




Figure 2. Optimizing inertial sorter module (A) PDMS device (B) Larger CTCs (green) go to innermost outlet channel. (C) Fluorescent beads, (D) fluorescent cells and (E) high-speed imaging demonstrate cancer cell sorting. Analysis of cell distribution in outlets both in PBS buffer (F) and blood (G).

Author

This article is protected by copyright. All rights reserved.




Figure 3. Integrated Device Operation (A) Magnetic Sorter outlets show highly enriched cancer cells separating from waste cells (B) Immunofluorescence staining used in identifying cancer cells (PANC-1) and WBCs with inset of magnetically labeled cancer cell (10 μm scale bar) (C) Recovery of PANC-1 cells spiked into whole blood at 100 and 1000 cells mL^{-1} . (D) Number of contaminating WBCs for 100 and 1000 PANC-1 cells spiked into whole blood.

Author Man




Figure 4. Processing PDAC samples (A) Cell counts of CTCs in 1mL of patient blood. Percent purities are indicated (B) Corresponding WBCs for the same samples (C) Log Exp values for 10 highest expressed miRNAs (D) Fluorescent and brightfield image of CTCs and (E) WBCs identified through immunofluorescent staining (10 μ m scale bar).

Author

This article is protected by copyright. All rights reserved.




Figure 5. Messenger RNA profiling of CTC enriched PDAC samples from 3 borderline (B) and 2 locally advanced (L) patients as well as 3 metastatic patients (M). (A) Heat map plot reveal that 2 metastatic profiles cluster together as a separate group from others excluding 1 outlier patient. Seven genes (green box) distinguish between the two groups. (B) Log fold increase plot show mRNAs increased in borderline versus metastatic patients. (* indicates samples with $p=0.036$).

This article is protected by copyright. All rights reserved.

Author Manuscript

WILEY-VCH

This article is protected by copyright. All rights reserved.



Keywords: cell sorting, immunomagnetism, microfluidics, microRNA, circulating tumor cells

R. M. Jack, M. M. G. Grafton, D. Rodrigues, M. D. Giraldez, C. Griffith, R. Cieslak, M. Zenali, C. Kumar-Sinha, E. Azizi, M. Wicha, M. Tewari, D. M. Simeone, and S. Nagrath*

The clinical potential of circulating tumor cells (CTCs) in managing cancer metastasis is significant. However low CTC isolation purities from patient blood has hindered sensitive molecular assays of these rare cells. Described herein, is the ultra-pure isolation of CTCs from patient blood samples and how this platform has enabled highly specific molecular (miRNA) profiling of patient CTCs.

Ultra-Specific Isolation of Circulating Tumor Cells Enables Rare-Cell Micro-RNA Profiling

Author Manuscript



CTC microRNA and mRNA profiling

Author M

REFERENCES:

1. Scheel, C.; Weinberg, R. A., Cancer Stem Cells and Epithelial–Mesenchymal Transition: Concepts and Molecular Links. *Seminars in Cancer Biology* **2012**.
2. Negin, B.; Cohen, S., Circulating Tumor Cells in Colorectal Cancer: Past, Present, and Future Challenges. *Current Treatment Options in Oncology* **2010**, *11*, 1-13.
3. de Albuquerque, A.; Kubisch, I.; Breier, G.; Stamminger, G.; Fersis, N.; Eichler, A.; Kaul, S.; Stolzel, U., Multimarker Gene Analysis of Circulating Tumor Cells in Pancreatic Cancer Patients: A Feasibility Study. *Oncology* **2012**, *82*, 3-10.
4. Yu, M.; Stott, S.; Toner, M.; Maheswaran, S.; Haber, D. A., Circulating Tumor Cells: Approaches to Isolation and Characterization. *The Journal of Cell Biology* **2011**, *192*, 373-382.
5. Nagrath, S., et al., Isolation of Rare Circulating Tumour Cells in Cancer Patients by Microchip Technology. *Nature* **2007**, *450*, 1235-1239.
6. Wang, S.; Wang, H.; Jiao, J.; Chen, K. J.; Owens, G. E.; Kamei, K. i.; Sun, J.; Sherman, D. J.; Behrenbruch, C. P.; Wu, H., Three-Dimensional Nanostructured Substrates toward Efficient Capture of Circulating Tumor Cells. *Angewandte Chemie* **2009**, *121*, 9132-9135.
7. Liu, H.; Liu, X.; Meng, J.; Zhang, P.; Yang, G.; Su, B.; Sun, K.; Chen, L.; Han, D.; Wang, S., Hydrophobic Interaction-Mediated Capture and Release of Cancer Cells on Thermoresponsive Nanostructured Surfaces. *Advanced Materials* **2013**, *25*, 922-927.
8. Di Carlo, D.; Edd, J. F.; Irimia, D.; Tompkins, R. G.; Toner, M., Equilibrium Separation and Filtration of Particles Using Differential Inertial Focusing. *Analytical Chemistry* **2008**, *80*, 2204-2211.
9. Hou, H. W.; Warkiani, M. E.; Khoo, B. L.; Li, Z. R.; Soo, R. A.; Tan, D. S.-W.; Lim, W.-T.; Han, J.; Bhagat, A. A. S.; Lim, C. T., Isolation and Retrieval of Circulating Tumor Cells Using Centrifugal Forces. *Scientific Reports* **2013**, *3*, 1259.
10. Bhagat, A. A. S.; Kuntaegowdanahalli, S. S.; Papautsky, I., Continuous Particle Separation in Spiral Microchannels Using Dean Flows and Differential Migration. *Lab on a Chip* **2008**, *8*.
11. Kuntaegowdanahalli, S. S.; Bhagat, A. A. S.; Kumar, G.; Papautsky, I., Inertial Microfluidics for Continuous Particle Separation in Spiral Microchannels. *Lab on a Chip* **2009**, *9*.
12. Ozkumur, E., et al., Inertial Focusing for Tumor Antigen–Dependent and –Independent Sorting of Rare Circulating Tumor Cells. *Science Translational Medicine* **2013**, *5*, 179ra47.
13. Liu, R. H.; Stremmler, M. A.; Sharp, K. V.; Olsen, M. G.; Santiago, J. G.; Adrian, R. J.; Aref, H.; Beebe, D. J., Passive Mixing in a Three-Dimensional Serpentine Microchannel. *Microelectromechanical Systems, Journal of* **2000**, *9*, 190-197.
14. Kim, K. S.; Park, J. K., Magnetic Force-Based Multiplexed Immunoassay Using Superparamagnetic Nanoparticles in Microfluidic Channel. *Lab Chip* **2005**, *5*, 657-64.

15. Gleghorn, J. P.; Pratt, E. D.; Denning, D.; Liu, H.; Bander, N. H.; Tagawa, S. T.; Nanus, D. M.; Giannakakou, P. A.; Kirby, B. J., Capture of Circulating Tumor Cells from Whole Blood of Prostate Cancer Patients Using Geometrically Enhanced Differential Immunocapture (Gedi) and a Prostate-Specific Antibody. *Lab on a Chip* **2010**, *10*, 27-29.
16. Stott, S. L., et al., Isolation of Circulating Tumor Cells Using a Microvortex-Generating Herringbone-Chip. *Proceedings of the National Academy of Sciences* **2010**.
17. Bloomston, M.; Frankel, W. L.; Petrocca, F.; Volinia, S.; Alder, H.; Hagan, J. P.; Liu, C. G.; Bhatt, D.; Taccioli, C.; Croce, C. M., MicroRNA Expression Patterns to Differentiate Pancreatic Adenocarcinoma from Normal Pancreas and Chronic Pancreatitis. *JAMA* **2007**, *297*, 1901-8.
18. Lopez-Casas, P. P.; Lopez-Fernandez, L. A., Gene-Expression Profiling in Pancreatic Cancer. *Expert review of molecular diagnostics* **2010**, *10*, 591-601.
19. Dillhoff, M.; Liu, J.; Frankel, W.; Croce, C.; Bloomston, M., MicroRNA-21 Is Overexpressed in Pancreatic Cancer and a Potential Predictor of Survival. *J Gastrointest Surg* **2008**, *12*, 2171-6.
20. Sadakari, Y.; Ohtsuka, T.; Ohuchida, K.; Tsutsumi, K.; Takahata, S.; Nakamura, M.; Mizumoto, K.; Tanaka, M., MicroRNA Expression Analyses in Preoperative Pancreatic Juice Samples of Pancreatic Ductal Adenocarcinoma. *JOP* **2010**, *11*, 587-92.
21. Buchholz, M.; Schatz, A.; Wagner, M.; Michl, P.; Linhart, T.; Adler, G.; Gress, T. M.; Ellenrieder, V., Overexpression of C-Myc in Pancreatic Cancer Caused by Ectopic Activation of Nfatc1 and the Ca²⁺/Calcineurin Signaling Pathway. *The EMBO journal* **2006**, *25*, 3714-24.
22. Bronsert, P., et al., Prognostic Significance of Zinc Finger E-Box Binding Homeobox 1 (Zeb1) Expression in Cancer Cells and Cancer-Associated Fibroblasts in Pancreatic Head Cancer. *Surgery* **2014**, *156*, 97-108.
23. Rowland, B. D.; Peeper, D. S., Klf4, P21 and Context-Dependent Opposing Forces in Cancer. *Nature reviews. Cancer* **2006**, *6*, 11-23.
24. Beger, C.; Ramadani, M.; Meyer, S.; Leder, G.; Kruger, M.; Welte, K.; Gansauge, F.; Beger, H. G., Down-Regulation of Brca1 in Chronic Pancreatitis and Sporadic Pancreatic Adenocarcinoma. *Clinical cancer research : an official journal of the American Association for Cancer Research* **2004**, *10*, 3780-7.
25. Chou, A., et al., Clinical and Molecular Characterization of Her2 Amplified-Pancreatic Cancer. *Genome Med* **2013**, *5*, 78.
26. Ting, D. T., et al., Single-Cell Rna Sequencing Identifies Extracellular Matrix Gene Expression by Pancreatic Circulating Tumor Cells. *Cell reports* **2014**, *8*, 1905-18.