

DR SAVINO SCIASCIA (Orcid ID : 0000-0003-1266-9441)

PROF. VITTORIO PENGO (Orcid ID : 0000-0003-2064-6071)

Article type : Original Article

Comparison of Real World and Core Laboratory Lupus Anticoagulant Results from the Antiphospholipid Syndrome (APS) ACTION Clinical Database and Repository

Authors:

Maria Efthymiou¹, Ian J. Mackie¹, Philip J. Lane¹, Danieli Andrade², Rohan Willis³, Doruk Erkan⁴, Savino Sciascia⁵, Steven Krillis⁶, Elisa Bison⁷, Margarete Borges Galhardo Vendramini², Zurina Romay-Penabad³, Miao Qi⁶, Maria Tektonidou⁸, Amaia Ugarte⁹, Cecilia Chighizola¹⁰, H. Michael Belmont¹¹, Maria Angeles Aguirre¹², Lanlan Ji¹³, D. Ware Branch¹⁴, Guilherme de Jesus¹⁵, Paul R. Fortin¹⁶, Laura Andreoli¹⁷, Michelle Petri¹⁸, Ricard Cervera¹⁹, Esther Rodriguez²⁰, Jason S. Knight²¹, Tatsuya Atsumi²², Joann Vega⁷, Ecem Sevim⁷, Maria Laura Bertolaccini²³, Vittorio Pengo⁶, Hannah Cohen^{1,24}; on behalf of APS ACTION*

Affiliations:

¹Haemostasis Research Unit, Department of Haematology, University College London, UK

²University of Sao Paulo, Sao Paulo, Brazil

³Antiphospholipid Standardization Laboratory, University of Texas Medical Branch, Galveston, TX, USA

⁴Barbara Volcker Center for Women and Rheumatic Diseases, Hospital for Special Surgery, Weill Cornell Medicine, New York, NY, USA

⁵Center of Research of Immunopathology and Rare Diseases, University of Turin, Italy

This is the author manuscript accepted for publication and has undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the [Version of Record](#). Please cite this article as [doi: 10.1111/JTH.14596](https://doi.org/10.1111/JTH.14596)

This article is protected by copyright. All rights reserved

⁶Department of Infectious Diseases, Immunology and Sexual Health, St George Hospital, Sydney, Australia

⁷University Hospital Padova, Padova, Italy

⁸ National and Kapodistrian University of Athens, Greece

⁹ Hospital Universitario Cruces, Internal Medicine, Barakaldo, País Vasco, Spain

¹⁰ University of Milan, Milan, Italy

¹¹New York University, School of Medicine, New York, USA

¹²Maimonides Institute for Biomedical Research of Cordoba, Cordoba, Spain

¹³Rheumatology and Immunology Department, Peking University First Hospital, Beijing, China

¹⁴University of Utah and Intermountain Healthcare, Salt Lake City, UT, USA

¹⁵Rio de Janeiro State University, Rio de Janeiro, Brazil

¹⁶CHU de Quebec – Université Laval, Quebec, Canada

¹⁷Rheumatology and Clinical Immunology, Spedali Civili; Department of Clinical and Experimental Sciences, University of Brescia, Brescia, Italy

¹⁸Johns Hopkins University School of Medicine, Department of Medicine, Baltimore, Maryland

¹⁹Department of Autoimmune Diseases, Hospital Clínic Institut d'Investigacions Biomèdiques August Pi i Sunyer (IDIBAPS), Barcelona, Catalonia, Spain

²⁰Hospital Universitario 12 de Octubre, Madrid, Spain

²¹Division of Rheumatology, University of Michigan, Michigan, USA

²²Hokkaido University Hospital, Sapporo, Japan

²³Academic Department of Vascular Surgery, School of Cardiovascular Medicine & Sciences, King's College London, London, UK

²⁴Department of Haematology, University College London Hospitals NHS Foundation Trust

Corresponding author: Maria Efthymiou

Haemostasis Research Unit, Department of Haematology, University College London, 1st Floor, 51 Chenies Mews, WC1E 6HX, London, UK;

Tel: 0044 207 679 6419/6420

Email: m.efthymiou@ucl.ac.uk

Abstract word count: 250 words

Manuscript word count: 4939

Number of Tables and Figures: 5 Tables and 3 Figures

Essentials

- Variability remains a challenge in lupus anticoagulant (LA) testing
- Good agreement in LA performance achieved by use of same reagents, analyser type and protocols
- At least some of the local/hospital laboratory LA results might have been unreliable in ~25% of non-anticoagulated and 23% of anticoagulated samples
- Standardised LA testing is required to underpin accurate interpretation of clinical data

Keywords: antiphospholipid antibodies, antiphospholipid syndrome, lupus anticoagulant, Russells viper venom time (RVVT), thrombosis

Running title: Lupus anticoagulant test validation in APS ACTION

Summary

Background: Variability remains a challenge in lupus anticoagulant (LA) testing.

Objective: To validate LA test performance between Antiphospholipid Syndrome Alliance for Clinical Trials and International Networking (APS ACTION) Core laboratories and examine agreement in LA status between Core and local/hospital laboratories contributing patients to this prospective registry.

Methods: Five Core laboratories used the same analyser, protocol, and characterised samples for LA validation. Non-anticoagulated registry samples were retested at the corresponding regional Core laboratories and anticoagulated samples at a single Core

laboratory. Categorical agreement and discrepancies in LA status between Core and local/hospital laboratories were analysed.

Results: Clotting times for the reference/characterised plasmas used for normalised ratios were similar between Core laboratories (CV <4%); precision and agreement for LA positive/negative plasma were similar (all CV ≤5%) in the four laboratories that completed both parts of the validation exercise. 418 registry samples underwent LA testing. Agreement for LA positive/negative status between Core and local/hospital laboratories was observed in 87% (115/132) non-anticoagulated and 77% (183/237) anticoagulated samples. However, 28.7% (120/418) of samples showed discordance between the Core and local/hospital laboratories or equivocal LA results. Some of the results of the local/hospital laboratories might have been unreliable in 24.7% (41/166) and 23% (58/252) of the total non-anticoagulated and anticoagulated samples, respectively. Equivocal results by the Core laboratory might have also contributed to discordance.

Conclusions: Laboratories can achieve good agreement in LA performance by use of same reagents, analyser type, and protocols. The standardised Core laboratory results underpin accurate interpretation of APS ACTION clinical data.

Introduction

Antiphospholipid syndrome (APS) is a potentially life-threatening acquired autoimmune disorder manifested by thrombosis (arterial, venous or microvascular) and/or obstetric morbidity in association with persistently positive antiphospholipid antibodies (aPL), i.e. lupus anticoagulant (LA), IgG and/or IgM anticardiolipin (aCL) and/or anti-beta 2 glycoprotein I antibodies (aβ2GPI) [1]. Accurate laboratory diagnosis of APS is crucial as these patients inherently differ from other patients with similar clinical features. In particular, the presence of aPL is associated with an increased risk of recurrent thrombosis and hence these patients are candidates for long term anticoagulation [2]. Detection of LA is problematic because of its heterogeneity and the variable sensitivity and specificity of different LA tests. False-positive or false-negative LA status may be detrimental to patient management as it could lead to unnecessary anticoagulation with the risk of bleeding, or inadequate thromboprophylaxis and risk of thrombosis recurrence, respectively. Detection of LA in plasmas from anticoagulated patients is especially problematic because of their prolonged clotting times. Recommendations published in 2009 by the International Society of Thrombosis and Haemostasis Scientific Standardisation Subcommittee (ISTH SSC) on Lupus Anticoagulant/Phospholipid Antibodies for the detection of LA, as well as the British Society of Haematology (BSH) (2012) and Clinical & Laboratory Standards Institute (CLSI) guidelines (2014) have proven to be useful guidance on how to perform LA tests. LA

detection should be a three step procedure, comprising a screening step with a LA sensitive phospholipid; a mixing step with normal plasma to demonstrate an inhibitor; and a confirmatory step to show phospholipid dependence [3-5].

The AntiPhospholipid Syndrome Alliance For Clinical Trials and InternatiOnal Networking (APS ACTION) is an international network created specifically to design and conduct well-designed, large-scale, multi-centre studies in persistently aPL-positive patients [6]. Since 2010, APS ACTION has been conducting a 10-year international clinical database and blood sample repository ('registry') to study the disease course in aPL-positive patients with/without other systemic autoimmune diseases. The network includes 50 physician-scientists from 25 international centres. APS ACTION registry laboratory inclusion criteria are moderate/high titre aCL and/or a β 2GPI and/or LA test based on the International consensus criteria [1], tested at least twice 12 weeks apart, within one year prior to enrolment. A secure web-based data capture system (REDCap) is used to store patient information. Blood samples and clinical information are collected from patients at inclusion (baseline) and then annually. APS ACTION has sought to standardize aPL laboratory testing that is of paramount importance to improve the interpretation and reliability of laboratory tests for associated clinical trials and research studies. Five APS ACTION Core laboratories were initially set up worldwide, in Sao Paulo (Brazil), Sydney (Australia), Galveston (USA), Padova (Italy), and London (UK), to store blood samples from hospitals in their continent/geographic region. Each Core laboratory performs aPL tests using standard protocols and reagents to confirm the original hospital result.

The aims of our study were to: a) validate the LA test performance between the five APS ACTION Core laboratories; and b) examine the degree of agreement in LA status between Core and local/hospital laboratories contributing patients to the registry.

Methods

Validation of Core laboratories LA test performance: All five Core laboratories used an ACL TOP500 analyser (Werfen, Barcelona, Spain) and performed the dilute Russell's viper venom time (DRVVT) and silica clotting time (SCT) with the same Lot numbers of HemosIL DRVVT Screen/Confirm and HemosIL Silica Clotting Time (SCT) (SCT Screen and SCT Confirm reagents). All laboratories were provided and followed a common study protocol and laboratory manual for LA testing designed specifically for APS ACTION. All laboratories used the same set of characterised plasmas for the validation exercise to establish agreement in

LA testing between the Core laboratories as follows: 1st International Reference Panel for LA (National Institute for Biological Standards and Control (NIBSC), South Mimms, UK; comprising: LA negative (NLA), moderate positive (MLA), and strong positive (SLA) LA samples), as well as HemosIL LA Negative (LA-) & HemosIL LA Positive (LA+) Control plasmas (QC), and these were tested on each of three working days. The manufacturer's cut off values for all tests were validated in each Core laboratory using plasmas from at least 40 healthy normal subjects prior to any tests being performed, according to national and international guidelines [4;7]. The results were calculated as normalised ratios, using local pooled normal plasma, or where this was not available, a commercial pool or LA negative control plasma. The within and between-centre precision was investigated for each plasma. The five labs were anonymised as A-E in no particular order.

Assessment of agreement in LA status between Core and local/hospital laboratories:

Each Core laboratory assessed LA status in samples collected from their allocated local/hospital laboratories within the first year of entry into the registry. All laboratories followed the APS ACTION protocol for preparation of samples, storage and LA testing. Samples were processed within four hours of collection and platelet poor plasma (PPP) was prepared from blood withdrawn by venepuncture in 0.109 M sodium citrate 9:1, then double-centrifuged at $1500 \times g$ for 15 minutes. To reduce variability, samples were stored at -80°C immediately after preparation and shipped on dry ice as this has been reported to cause minimum interference with LA testing and analysis [8;9]. Samples were thawed uncapped at 37°C (using a water bath) prior to assessment as recommended by CLSI [5]. Samples from non-anticoagulated patients (n=166) were assessed using DRVVT and SCT reagents, as detailed above, and an ACL TOP500 analyser. Equal volume mixtures of patient and normal plasma were tested to confirm the presence of an inhibitor.

LA status for patients receiving anticoagulation was assessed at a single Core laboratory (UK). For patients receiving vitamin K antagonists (VKA), the DRVVT was performed with screen and confirm reagents on equal volume mixtures of patient/normal plasma (which if positive confirmed the presence of an inhibitor and phospholipid dependence); and the Taipan venom time (TVT)/Ecarin clotting time (ECT; Diagnostic Reagents Ltd, Thame, UK) ratio was performed on undiluted patient samples as well as on equal volume mixtures of patient/normal plasma if the TVT was prolonged. The SCT screen and confirm tests on equal volumes of patient/normal plasma was not performed in patients receiving VKA due to the limited volume of plasma available. Patients receiving low molecular weight heparin (LMWH) prophylaxis were assessed by DRVVT (Screen/Confirm) and SCT (DRVVT reagents are insensitive to heparin or LMWH up to 1.0IU/ml; SCT reagents are insensitive to heparin up to 0.5IU/ml and LMWH to 1.0IU/ml). For patients receiving rivaroxaban, a direct

oral anticoagulant (DOAC) factor (F) Xa inhibitor, LA status was assessed by TVT/ECT ratio and rivaroxaban levels were measured with an anti-FXa amidolytic assay using specific calibrators.

LA status was considered positive if the patient/normal DRVVT or SCT screen ratio was ≥ 1.20 and the screen/confirm normalised ratio was ≥ 1.20 . TVT/ECT was considered positive if the TVT was prolonged and the normalised TVT/ECT ratio was > 1.20 . The TVT/ECT ratio cut off of 1.20 [10] was established locally in accordance with CLSI guidelines [7]

In all tests, evidence of inhibition was provided by testing equal volume mixtures of patient and normal plasma. Mixing tests were interpreted as indicating an inhibitor if the test/normal ratio was greater than the cut-off for normal plasma. Results were recorded as equivocal if a mixing test was not performed (due to insufficient plasma volume) or if an inhibitor was not demonstrated, and the test/confirm ratio was positive. Results were reported as positive, negative (“not detected” in the case of anticoagulated patients), or equivocal (where there was no evidence of an inhibitor or suspicion of an underlying coagulopathy).

Statistical Analysis: Agreement of categorical positive and negative aPL test results was assessed using κ -coefficients (< 0.20 , poor; $0.21-0.40$ fair; $0.41-0.60$ moderate; $0.61-0.80$ good; $0.81-1.00$ very good) [11] and with the Holley and Gilford's G test with a 99% confidence as our results were skewed. The G-test is equivalent to several other inter-rater agreement statistics that have been proposed over the years. In contrast to κ -Cohen, the calculation of chance agreement in G does not depend on the obtained frequencies but is defined a priori. The G-test statistic is a measure of how much overall variation there is from an ideal prediction that you would expect if all versions were the same. Even if there is no hard rule, however a common benchmark is 99% confidence, meaning that with the obtained G-test values, we have less than a 1% chance of making a mistake at the point in time where we made our decision on test agreement [12]. All statistical analysis was performed using SPSS 19.0 (IBM Software, NY).

Results

Evaluation of LA test performance

Two Core laboratories (B, C) used HemosIL Normal Control Plasma (Werfen) for the calculation of normalised ratios, while three (A, D, E) used in-house or commercial lyophilised (Pathway Diagnostics, Dorking, UK) or frozen (Cryocheck, Precision Biologics,

Dartmouth, Canada) pooled normal plasma. These all yielded very similar clotting times for DRVVT Screen (32.7-34.2s) and Confirm (30.5-32.4s), with between laboratory CV values <3%; and for SCT Screen (34.7-35.3s) and Confirm (35.4-36.0s), CV<4%.

Precision and agreement were generally good between all Core laboratories for LA negative control and test plasmas using both DRVVT Screen and Confirm reagents. The between day coefficient of variation (CV) of clotting time was <2.5% for LA negative control plasma and <5.0% for LA positive plasmas (i.e. MLA & SLA). However, one laboratory (D) tended to show longer clotting times and higher ratios for LA positive samples (MLA & SLA) with DRVVT Screen reagent (Fig 1), but not with Confirm reagent. Similar disparity between laboratories in clotting times and ratios was observed in mixing tests (data not shown). Despite both laboratory D, and to a lesser extent laboratory B, obtaining higher average DRVVT Screen/Confirmed normalised ratios from the other Core laboratories, all laboratories correctly identified the LA status of the LA+ QC, NLA, MLA, and SLA samples (average values over 3 days shown in Table 1).

Four Core laboratories returned results for SCT (Laboratories A, B, C, E) and showed good agreement for clotting times, with all CV values <3.2%. Although test/normal ratios for SCT Screen reagent showed some variability (Fig 2), these discrepancies disappeared when normalised screen/confirm ratios were calculated (Table 1). Laboratory D played no further role as a Core laboratory for LA assessments and did not contribute data to the results sections below.

Agreement in LA status between Core and local/hospital laboratories

Samples from 508 APS ACTION patients were analysed from a total of 573 registered (until June 2015) (Fig 3). Ninety samples (10 non-anticoagulated and 80 anticoagulated) were excluded from the analysis, with the reasons for exclusion due to: insufficient volume of citrate plasma for further analysis (n=38); the local/hospital laboratory submitting only aCL/a β 2GPI results; or LA status not being available (n=52). Samples from the remaining 418 patients (166 non-anticoagulated and 252 anticoagulated) were analysed (Fig 3).

Of the 418 patients analysed, 222 (53.1%) were diagnosed with primary APS (according to Sapporo criteria [1] 58 (13.9%) with aPL (but insufficient Sapporo criteria to fulfil APS), 86 (20.6%) as APS with other systemic autoimmune disease, and 52 (12.4%) had aPL without Sapporo APS criteria, but with other systemic autoimmune disease. Of the 418 patients, 218 (52.2%) were classified as thrombotic (either venous or arterial or both), 50 (12.0%) as thrombotic with pregnancy morbidity, 40 (9.6%) as obstetrical APS and 110 were aPL positive without APS (26.3%).

For the non-anticoagulated patient samples (n=166) analysed, local/hospital laboratories submitted LA results (for the year prior to data entry): on more than three occasions in 92 patients (55.4%); two occasions in 59 patients (31.3%); and one occasion in 15 patients (9.0%). Of the 252 anticoagulated patient samples analysed, LA status was assessed and reported by the local/hospital laboratories in more than three hospital visits in 154 patients (61.1%), two visits in 63 patients (25.0%), one visit in 35 patients (13.9%). Local/hospital laboratories that submitted LA results on only one occasion for the year prior to data entry had submitted either aCL or a β 2GPI results on more than two occasions (thus criteria for APS diagnosis had been fulfilled) before inclusion in the database.

Local/hospital laboratories were asked to record the method of LA assessment for each occasion tested (in the year prior to study entry). A total of 1039 tests were reported: 5.1% performed activated partial thromboplastin time (aPTT)-based tests, 31.1% DRVVT, 53.5% aPTT-based tests and DRVVT, 8.3% reported "unknown" and 2% as "other" (rather than DRVVT or APTT).

At inclusion in the APS ACTION registry, LA status (positive/negative) reported by the local/hospital laboratories was as follows: for non-anticoagulated samples: 70% (111/166) LA positive and 30.0% (50/166) LA negative and for anticoagulated samples: 84% (211/252) LA positive and 16% (41/252) LA negative.

Non-anticoagulated patient samples

For the 166 non-anticoagulated patients, 132 were recorded as LA positive or negative and amongst these there was 87.1% agreement between the Core and the local/hospital laboratories about LA classification (K coefficient= 0.589, $p < 0.001$; G test statistic 32.6 >99% confidence, Table 2). Seventeen samples (12.9%) were discordant between the two with twelve samples identified as LA negative in the local/hospital laboratory, but positive in the Core laboratories, while five samples showed discordance in the opposite way (Table 2).

We examined possible reasons for the discordance in LA status between Core and local/hospital laboratories in these 17 patient samples. Four were only tested on one occasion at the local/hospital laboratory; six were tested more than once, but the LA status varied; in three the LA method was not specified, or only one LA method was performed (since two were recorded as LA negative, there could have been under detection of LA); Only four samples out of the 17 discordant were tested on more than one occasion with consistent results (1 positive and 3 negative) (Table 3).

The results in 34 further samples (20.5%, 34/166) were considered to be equivocal by the Core laboratories because an inhibitor could not be demonstrated. Of these, 21 had been

recorded as negative and 13 as positive by the local/hospital laboratory (Table 3). Ten of these patients only had a LA assessment on one occasion prior to inclusion in the registry. Another ten were tested more than once at the local/hospital laboratory, but they obtained variable results. In eight patients, either the method of LA assessment was not specified or only one LA method was performed (since five of these were recorded as LA negative, there could have been under-detection of LA). In only six out of the 34 discordant samples, LA status was tested on more than one occasion and consistent LA results were recorded (three as LA positive and three as negative) (Table 3). On this basis, it could be argued that the local/hospital laboratory determination of LA might not be reliable in at least some of the 13/17 discordant samples and 28/34 samples that gave equivocal results in the Core laboratories. This would potentially account for 80.4% (41/51) of samples where there was a disagreement in LA status between the Core and local/hospital laboratories i.e. 41/166 (24.7%) of non-anticoagulated samples. Results from the Core laboratories might also have contributed to the discordance, especially the ones judged equivocal i.e. 20.5%, 34/166 of non-anticoagulated samples.

Anticoagulated patient samples

All samples from anticoagulated patients were tested at a single Core laboratory (London, UK), which had the appropriate test repertoire available. The majority of patients receiving anticoagulants were on long term warfarin or other VKA (n=224), six were receiving the DOAC FXa inhibitor rivaroxaban, 18 LMWH, and four both VKA and LMWH. Of the 252 anticoagulated patient samples, 15 were considered to be equivocal by the Core laboratory, since they were negative by DRVVT and an inhibitor could not be demonstrated in the TVT. The remaining samples showed 77.2% (183/237) agreement in LA classification. Discordance between the Core and local/hospital laboratories was identified in 22.7% (54/237) of samples with equal numbers of samples (27 each) identified as negative by either the Core or local/hospital laboratories and positive by the other (Table 4).

There was an additional 6.0% (15/252) of patients showing equivocal results; three were tested on only one occasion at the local/hospital laboratory; four were tested more than once, but with variable results; in five the LA methods were not specified or only one LA method was used; only three were tested more than once with consistent results (two LA positive, one negative) (Table 5).

We also examined possible reasons for the discordance in LA status between Core and local/hospital laboratories in the 54 anticoagulated samples, which were as follows: 15 were tested on only one occasion; 20 were tested more than once, but LA status varied; and in 11 the LA method was not specified or only one method was used. Only eight of the 54 samples

were tested on more than one occasion with consistent results (four LA positive and four negative) (Table 5). Of the patients with discordant results, the LA method at the local/hospital laboratory was recorded as “other” (than aPTT or DRVVT) in 10 cases, “aPTT-based” in five, as “unknown” in five, DRVVT in 14 and aPTT plus DRVVT in 20. On this basis, one could argue that the LA status reported by the local/hospital laboratories might not be reliable in at least some of the 84% (58/69) samples with discordant or equivocal LA status (46/54 discordant and 12/15 samples giving equivocal results), i.e. in 58/252 (23%) of the overall anticoagulated patients. Results judged equivocal by the Core laboratories (6.0%, 15/252) might also have contributed to the discordance observed with local/hospital laboratories in anticoagulated samples.

Discussion

In the first part of this prospective APS ACTION study, we demonstrated that good agreement and low variability in LA performance between the APS ACTION Core laboratories was achieved by use of the same reagents, analyser type, and protocols. Furthermore, in the second part of the study we established that LA status results between Core and local/hospital laboratories were concordant in 87% (115/132 non-anticoagulated) and 77% (anticoagulated) samples, when equivocal LA results were excluded. Importantly, however, we identified that 28.7% (120/418) of the total samples showed discordance or equivocal LA results between the Core and local/hospital laboratories. Thirteen percent (17/132) of non-anticoagulated and 22.7% (54/237) of anticoagulated patient samples were discordant between the two laboratories (when equivocal results were excluded). A further 20.5% (34/166) and 6% (15/252) of non-anticoagulated and anticoagulated patient samples respectively, were identified as equivocal by the Core laboratories, but as positive or negative by the local/hospital laboratory. Further analysis indicated that the LA results in 80.4% (41/51) of non-anticoagulated and 84% (58/69) of anticoagulated discordant/equivocal samples may not be reliable. This accounted for 24.7% (of 166) of the total non-anticoagulated and 23% (of 252) anticoagulated samples respectively. These results indicate that it is important to use validated and standardised LA results, as provided by the Core laboratories, to underpin the accurate interpretation of APS ACTION clinical data.

There are many differences between haemostasis laboratories in the selection of LA tests, source of reagents, methodological detail, and results [2;13-16]. External quality assessment studies in Europe have shown considerable variability between laboratories for LA detection, particularly in samples with “weak” LA, with false negative and false positive rates of 10-20%

[17;18]. In North America, false negative rates up to 28% and false positive rates of around 11% have been reported, while in Australasian studies, false negative rates up to 50% and false positive rates of about 10% were observed [14;19]. These discrepancies appear to be due to a variety of pre- and post-analytical factors as well as performance of the tests. Within method, but not between method, agreement could be improved by the use of common normal plasma for ratio calculations and the adoption of recommendations from National and International guidelines on LA tests. Between and within method agreement can also be improved by use of the mean reference range for clotting time for ratio calculation [13;18;19].

The sensitivity and specificity of results also appears to depend on the potency of the LA and the method used. APTT based methods generally appear to be more sensitive to weak and moderate potency LA than DRVVT methods, but have lower specificity [14]. In one study, 29 different hospitals referred plasma samples from patients with lupus anticoagulant to a reference laboratory [8], where they were re-tested using DRVVT and kaolin clotting time. 24% of samples were LA negative on re-testing, either due to normal clotting times in screening tests, or complete correction in mixing tests to demonstrate an inhibitor. However, in this study, hospitals submitting samples had been asked to filter the plasma samples, which can introduce variability due to potential changes in clotting factor levels as well as inducing microparticle formation and consequent changes in the clotting time after freezing and thawing [20]. Sample preparation is critical for LA detection, as platelet contamination of plasma samples can lead to false negative LA results after the plasma is frozen and stored [9]. Local cut-off and reference ranges are important, as there are differences depending on the reagent and analyser used [21] as well as the method for calculating the values [22]. In our study, following venepuncture, blood samples were double centrifuged to obtain platelet poor plasma and stored at -80°C, in accordance with national and international guidelines [3-5].

Notwithstanding the good agreement and precision in LA testing demonstrated between four Core Laboratories in the first part of our study, it is clearly important to monitor laboratory performance, particularly if laboratories are not experienced in performing haemostasis tests. One Core laboratory (whose expertise is in solid phase aPL and other immunoassays), produced markedly different results from the other four and therefore took no further part in LA assessment for APS ACTION. In hospital laboratories that regularly perform LA tests, such discrepant performance would be identified through external quality assessment participation schemes such as external quality assurance services (EQAS), United Kingdom National External Quality Assessment Service (UK NEQAS), External quality Control of diagnostic Assays and Tests (ECAT foundation).

There are a variety of possible reasons for the observed discordance between Core and local/hospital laboratories in the second part of the study. Laboratories implement the various guidance documents [3-5] for LA detection in different ways and adherence is influenced by the type of automation available, local analyser protocols, laboratory computer systems and the economics of batch testing compared to single sample analysis. The guidelines recommend the performance of two different LA tests employing different methodologies, only one of which needs to indicate LA. However, in some laboratories, only one test system is available. This is a particular problem in samples from anticoagulated patients, where some test systems (e.g. aPTT based tests) may not be appropriate due to their sensitivity to warfarin [23], or where a heparin neutralising agent is not included in the reagent. According to APS criteria, if LA positivity is the sole laboratory criteria for diagnosis, it should be demonstrated on at least two occasions 12 weeks or more apart [1]. Some local/hospital laboratories contributing to APS ACTION only tested for LA on one occasion while they confirmed persistence of aCL/a β 2GPI, but not LA, more than once; other laboratories only used a single LA reagent; and some obtained variable results on different blood samples, meaning that LA may have been present but undetected in some patients. Our data suggesting that at least some of the local/hospital LA results may not be reliable in up to 24.7% of the total non-anticoagulated and 23% of the total anticoagulated samples, is of particular clinical importance as the result could have a major impact on patient management. Core laboratory results were limited because samples could only be tested at a single time point (and in some cases with different reagents and methods to local/hospital laboratories) and this could have introduced differences and contributed to the discordance with the local/hospital LA results. This is particularly critical for equivocal samples (20.5% and 6% of non-anticoagulated and anticoagulated samples, respectively) and these need later re-evaluation; longitudinal testing of registry samples collected in subsequent years will enable determination of the LA status of all patients in the APS ACTION registry. However, the purpose of the Core laboratories testing was to provide a highly standardised, consistent assessment of LA status.

International APS criteria and LA testing guidance also state that part of LA testing should be to demonstrate an inhibitor, usually achieved by testing mixtures of patient and normal plasmas [3-5]. The results for some samples were deemed to be equivocal in the Core laboratories, because a prolonged clotting time could not be demonstrated in plasma mixtures. However, this does not prove that they are LA negative, as aPL may have been diluted in the mixing test to below the threshold for detection; nor does a prolonged screening test but normal mixing test mean that the patient definitely has a "weak" LA, as the clotting time could have been prolonged for other reasons.

LA assessment in samples from patients receiving VKA anticoagulation is particularly challenging, since most LA tests require the presence of vitamin K dependent coagulation factors. As also concluded in our study, the discordance in LA status between the local/hospital and Core laboratories was more prevalent in the anticoagulated samples (22.7%, 54/237) when compared to non-anticoagulated ones (12.9%, 17/132) ($p=0.03$). LA testing in anticoagulated samples is generally discouraged as the probability for false positive/negative results is high [24;25]. However, determination of LA status is required in certain clinical settings as well as for full characterisation of aPL status of patients in research studies and registries such as APS ACTION. Where the INR is <4.0 , it is generally accepted that DRVVT screen and confirm procedures can be used on equal volume mixtures of patient and control plasma (with evidence of an inhibitor provided where the DRVVT screen normalised ratio is still abnormal) [26] APTT-based tests, including SCT, can be problematic, due to their vitamin K factor sensitivity and great prolongation of clotting time. Textarin or Taipan venoms, which are direct activators of prothrombin and are little affected by VKA, can be used, although textarin venom is currently difficult to obtain. Coupled with Ecarin venom, the TVT/ECT ratio is useful for LA testing in VKA anticoagulated patients [27-30]. The TVT/ECT ratio is less sensitive than the DRVVT, and thus a positive result may be useful, but a negative result should be considered as “not detected”. Taipan venom also shows low sensitivity to heparin pentasaccharide, although in the presence of long chain heparin antithrombin can inhibit the taipan venom by a template dependent mechanism similar to that of thrombin inhibition [31]. There is an even greater problem with samples from patients receiving DOACs. In our study, only six patients were receiving rivaroxaban, however, the number of patients receiving DOACs may increase in APS ACTION in future years. FXa inhibitors interfere with APTT-based tests and DRVVT and may thus result in erroneous results. It has been suggested that the DRVVT can be used in the drug trough period, however false positive results have been reported even at very low ($<50\text{ng/mL}$) rivaroxaban levels [32]. We assessed samples containing rivaroxaban using the TVT/ECT, which is not affected by rivaroxaban [33;34]. Both the TVT and ECT tests were performed with patient plasma rather than equal volume mixtures of patient and normal plasma, as the venom reagents lead to direct activation of factor II and show little effect of VKA and direct FXa inhibitor anticoagulation. However, a mixing test was performed for TVT in this study if the TVT clotting time was prolonged and if sufficient plasma was available, to establish the presence of an inhibitor. Limitations to this approach were that: only one LA test (TVT/ECT) was available for samples from DOAC-treated patients and performing a single test might increase the risk of false-negative results; samples were tested on one occasion only; and the sensitivity of the TVT/ECT is generally less than that for DRVVT [35]. This problem may be resolved in the future by the use of DOAC removal agents, which are

showing encouraging results [36;37] and with testing of longitudinal samples. The use of cut offs for undiluted plasma for interpretation of the mixing tests is also a limiting factor in this study as this might lead to reduced detection rates and could have increased the numbers of equivocal and false-negative results.

Conclusions

Good agreement in LA performance can be achieved, as in the APS ACTION Core Laboratories, by use of the same reagents, analyser type, and protocols. The observed agreement in LA test results (positive/negative) between Core and local/hospital laboratories of 87% non-anticoagulated samples and 77% anticoagulated samples (when equivocal results were excluded), appeared acceptable on initial review, especially when considering high inter-laboratory variability, and the difficulties of detecting LA in samples from anticoagulated patients as well as lack of adherence to guidelines. However, 28.7% (120/418) of samples showed discordance between the Core and local/hospital laboratories, or equivocal LA results. Analysis of possible reasons for discordance/equivocal LA suggested that at least some of the results of the local/hospital laboratories might not have been reliable in 24.7% (41/166) and 23% (58/252) of the total non-anticoagulated and anticoagulated samples, respectively. Results by the Core laboratories judged as equivocal might also have contributed in the discordance in 20.5% (34/166) and 6% (15/252) of non-anticoagulated and anticoagulated samples, respectively. Thus, standardised Core laboratory LA testing is required to underpin the accurate interpretation of APS ACTION clinical data. Studies on LA testing on the annual follow up samples from the APS ACTION registry patients will enable future re-evaluation of longitudinal LA data and enable correct identification and validation of LA status. Our study has highlighted the need for widespread standardisation of LA testing, to underpin accurate diagnosis and management of APS patients.

Addendum

Maria Efthymiou designed and carried out the study, analysed data, performed statistical analysis, and wrote the first draft of the manuscript. Ian J Mackie designed the study, analysed data and reviewed the manuscript. Philip J Lane performed experiments, data handling, and reviewed the manuscript. Hannah Cohen designed the study and provided critical revision of the manuscript. Vittorio Pengo, Danieli Andrade, Rohan Willis, Steve Krilis and Laura M Bertolaccini contributed to writing the laboratory manual for Core laboratories

and oversaw the laboratory work at their sites. Doruk Erkan was involved in designing the study and led the overall APS ACTION work. Savino Sciascia performed statistical analysis, Elisa Bison, Margarete Borges Galhardo Vendramini, Zurina Romay-Penabad, Miao Qi, performed experiments at the respective Core laboratory. Joann Vega and Ecem Sevim performed research duties including patient recruitment and data monitoring and maintenance of the APS ACTION database. Amaia Ugarte, Cecilia B. Chighizola, H. Michael Belmont, Lanlan Ji, M^a Angeles Aguirre Zamorano, Guilherme de Jesus, D. Ware Branch, Paul R. Fortin, Laura Andreoli, Michelle Petri, Esther Rodriguez, Ricard Cervera, Jason S. Knight, and Tatsuya Atsumi were involved in patient recruitment and data entry into the APS ACTION database and registry. All authors contributed to revising the intellectual content of the manuscript and final approval of the version to be published.

Conflict of interest

The authors declare that they have no conflict of interest. Prof Michelle Petri received funding support for the Hopkins Lupus Cohort by NIH RO1 AR069572

Acknowledgments

We would like to thank NIBSC, South Mimms, UK for kindly providing the 1st International Reference Panel for LA plasma; Werfen for providing the ACL TOP500 analyser and LA reagents; and Diagnostic Reagents Ltd for providing Taipan and Ecarin venoms; without having any input in the study and/or analysis of the data or results. Data management was performed using REDCAP provided by the Clinical and Translational Science Center at Weill Cornell Medicine (CTSC grant UL1 TR000457).

Reference List

- 1 Miyakis S, Lockshin MD, Atsumi T, Branch DW, Brey RL, Cervera R, Derksen RH, DE Groot PG, Koike T, Meroni PL, Reber G, Shoenfeld Y, Tincani A, Vlachoyiannopoulos PG, Krilis SA. International consensus statement on an update of the classification criteria for definite antiphospholipid syndrome (APS). *J Thromb Haemost* 2006; **4**: 295-306.

- 2 Tripodi A. Laboratory testing for lupus anticoagulants: a review of issues affecting results. *Clin Chem* 2007; **53**: 1629-35.
- 3 Pengo V, Tripodi A, Reber G, Rand JH, Ortel TL, Galli M, DE Groot PG. Update of the guidelines for lupus anticoagulant detection. Subcommittee on Lupus Anticoagulant/Antiphospholipid Antibody of the Scientific and Standardisation Committee of the International Society on Thrombosis and Haemostasis. *J Thromb Haemost* 2009; **7**: 1737-40.
- 4 Keeling D, Mackie I, Moore GW, Greer IA, Greaves M. Guidelines on the investigation and management of antiphospholipid syndrome. *Br J Haematol* 2012; **157**: 47-58.
- 5 CLSI Guideline. Laboratory Testing for the Lupus Anticoagulant; Approved Guideline. CLSI document H60-A, 2014. 2014. Clinical and Laboratory Standards Institute, Wayne, PA, USA.
- 6 Erkan D, Lockshin MD. APS ACTION--AntiPhospholipid Syndrome Alliance For Clinical Trials and InternatiOnal Networking. *Lupus* 2012; **21**: 695-8.
- 7 CLSI Guideline. Defining, Establishing, and Verifying Reference Intervals in the Clinical Laboratory; Approved Guideline. CLSI document EP28-A3c. 2008. 2008. Clinical and laboratory standard institute, Wayne, PA, USA
- 8 Pengo V, Biasiolo A, Gresele P, Marongiu F, Erba N, Veschi F, Ghirarduzzi A, de CE, Montaruli B, Testa S, Barcellona D, Tripodi A. Survey of lupus anticoagulant diagnosis by central evaluation of positive plasma samples. *J Thromb Haemost* 2007; **5**: 925-30.
- 9 Odsaeter IH, Lian IA, Bratberg K, Mikkelsen G. Dry ice exposure of plasma samples influences pH and lupus anticoagulant analysis. *Clin Chem Lab Med* 2015; **53**: 809-13.
- 10 Arachchilage DR, Mackie IJ, Efthymiou M, Isenberg DA, Machin SJ, Cohen H. Interactions between rivaroxaban and antiphospholipid antibodies in thrombotic antiphospholipid syndrome. *J Thromb Haemost* 2015; **10**.
- 11 Cohen J. Weighted kappa: nominal scale agreement with provision for scaled disagreement or partial credit. *Psychol Bull* 1968; **70**: 213-20.
- 12 Xu S, Lorber MF. Interrater agreement statistics with skewed data: evaluation of alternatives to Cohen's kappa. *J Consult Clin Psychol* 2014; **82**: 1219-27.

- 13 Adams M. Measurement of lupus anticoagulants: an update on quality in laboratory testing. *Semin Thromb Hemost* 2013; **39**: 267-71.
- 14 Dembitzer FR, Ledford Kraemer MR, Meijer P, Peerschke EI. Lupus anticoagulant testing: performance and practices by north american clinical laboratories. *Am J Clin Pathol* 2010; **134**: 764-73.
- 15 Devreese K, Hoylaerts MF. Laboratory diagnosis of the antiphospholipid syndrome: a plethora of obstacles to overcome. *Eur J Haematol* 2009; **83**: 1-16.
- 16 Moffat KA, Ledford-Kraemer MR, Plumhoff EA, McKay H, Nichols WL, Meijer P, Hayward CP. Are laboratories following published recommendations for lupus anticoagulant testing? An international evaluation of practices. *Thromb Haemost* 2009; **101**: 178-84.
- 17 Jennings I, Kitchen S, Woods TA, Preston FE, Greaves M. Potentially clinically important inaccuracies in testing for the lupus anticoagulant: an analysis of results from three surveys of the UK National External Quality Assessment Scheme (NEQAS) for Blood Coagulation. *Thromb Haemost* 1997; **77**: 934-7.
- 18 Jennings I, Greaves M, Mackie IJ, Kitchen S, Woods TA, Preston FE. Lupus anticoagulant testing: improvements in performance in a UK NEQAS proficiency testing exercise after dissemination of national guidelines on laboratory methods. *Br J Haematol* 2002; **119**: 364-9.
- 19 Favaloro EJ, Bonar R, Marsden K. Internal quality control and external quality assurance in testing for antiphospholipid antibodies: Part II--Lupus anticoagulant. *Semin Thromb Hemost* 2012; **38**: 404-11.
- 20 Favaloro EJ. Preanalytical variables in coagulation testing. *Blood Coagul Fibrinolysis* 2007; **18**: 86-9.
- 21 Gardiner C, Mackie IJ, Malia RG, Jones DW, Winter M, Leeming D, Taberner DA, Machin SJ, Greaves M. The importance of locally derived reference ranges and standardized calculation of dilute Russell's viper venom time results in screening for lupus anticoagulant. *Br J Haematol* 2000; **111**: 1230-5.
- 22 Tripodi A, Chantarangkul V, Cini M, Devreese K, Dlott JS, Giacomello R, Gray E, Legnani C, Martinuzzo ME, Pradella P, Siegemund A, Subramanian S, Suchon P,

- Testa S. Variability of cut-off values for the detection of lupus anticoagulants: results of an international multicenter multiplatform study. *J Thromb Haemost* 2017; **15**: 1180-90.
- 23 Dembitzer FR, Suarez Y, Aledort LM, Peerschke EI. Screening coagulation testing using the APTT: which reagent to choose? *Am J Hematol* 2010; **85**: 726.
- 24 Aboud M, Morel-Kopp MC, Ward C, Coyle L. False-negative or false-positive: laboratory diagnosis of lupus anticoagulant at the time of commencement of anticoagulant. *J Thromb Haemost* 2010; **8**: 2070-3.
- 25 Tripodi A, Moia M, Pengo V. False-negative or false-positive: laboratory diagnosis of lupus anticoagulant at the time of commencement of anticoagulant: a rebuttal. *J Thromb Haemost* 2011; **9**: 1435-6.
- 26 Pengo V, Tripodi A, Reber G, Rand JH, Ortel TL, Galli M, DE Groot PG. Update of the guidelines for lupus anticoagulant detection. Subcommittee on Lupus Anticoagulant/Antiphospholipid Antibody of the Scientific and Standardisation Committee of the International Society on Thrombosis and Haemostasis. *J Thromb Haemost* 2009; **7**: 1737-40.
- 27 Rooney AM, McNally T, Mackie IJ, Machin SJ. The Taipan snake venom time: a new test for lupus anticoagulant. *J Clin Pathol* 1994; **47**: 497-501.
- 28 Parmar K, Lefkou E, Doughty H, Connor P, Hunt BJ. The utility of the Taipan snake venom assay in assessing lupus anticoagulant status in individuals receiving or not receiving an oral vitamin K antagonist. *Blood Coagul Fibrinolysis* 2009; **20**: 271-5.
- 29 Moore GW, Smith MP, Savidge GF. The Ecarin time is an improved confirmatory test for the Taipan snake venom time in warfarinized patients with lupus anticoagulants. *Blood Coagul Fibrinolysis* 2003; **14**: 307-12.
- 30 Triplett DA, Stocker KF, Unger GA, Barna LK. The Textarin/Ecarin ratio: a confirmatory test for lupus anticoagulants. *Thromb Haemost* 1993; **20**; **70**: 925-31.
- 31 Chen L, Rezaie AR. Proexosite-1-dependent recognition and activation of prothrombin by taipan venom. *J Biol Chem* 2004; **279**: 17869-74.
- 32 Martinuzzo ME, Forastiero R, Duboscq C, Barrera LH, Lopez MS, Ceresetto J, Penschsky D, Oyhamburu J. False-positive lupus anticoagulant results by DRVVT in

the presence of rivaroxaban even at low plasma concentrations. *Int J Lab Hematol* 2018; 10.

- 33 Arachchillage DR, Efthymiou M, Mackie IJ, Lawrie AS, Machin SJ, Cohen H. Rivaroxaban and warfarin achieve effective anticoagulation, as assessed by inhibition of TG and in-vivo markers of coagulation activation, in patients with venous thromboembolism. *Thromb Res* 2015; **135**: 388-93.
- 34 van Os GM, de LB, Kamphuisen PW, Meijers JC, DE Groot PG. Detection of lupus anticoagulant in the presence of rivaroxaban using Taipan snake venom time. *J Thromb Haemost* 2011; **9**: 1657-9.
- 35 Moore GW, Culhane AP, Maloney JC, Archer RA, Breen KA, Hunt BJ. Taipan snake venom time coupled with ecarin time enhances lupus anticoagulant detection in nonanticoagulated patients. *Blood Coagul Fibrinolysis* 2016; **27**: 477-80.
- 36 Favresse J, Lardinois B, Sabor L, Devalet B, Vandepapeliere J, Braibant M, Lessire S, Chatelain B, Jacqmin H, Douxfils J, Mullier F. Evaluation of the DOAC-Stop procedure to overcome the effect of DOACs on several thrombophilia screening tests. *Thromb Haemost* 2018; **2**: e202-e209.
- 37 Exner T, Michalopoulos N, Pearce J, Xavier R, Ahuja M. Simple method for removing DOACs from plasma samples. *Thromb Res* 2018; **163**:117-122.

Author Manuscript

Tables

Table 1: Normalised Screen/Confirm ratios in each Core laboratory for DRVVT and SCT: All Core laboratories identified correctly the LA status of the LA+ QC, NLA, MLA, and SLA samples (average values over three days \pm standard deviation are shown, normal cut-off taken as 1.20). Laboratory D did not return results for SCT. The LA-negative QC plasma results are not included as it mirrored the results indicated by the NLA.

DRVVT	LA+QC	NLA	MLA	SLA
Expected				
LA status:	Positive	Negative	Moderate	Strong
A	1.49 \pm 0.02	0.90 \pm 0.02	1.31 \pm 0.02	1.51 \pm 0.02
B	1.75 \pm 0.04	1.08 \pm 0.01	1.53 \pm 0.06	1.78 \pm 0.05
C	1.58 \pm 0.05	1.01 \pm 0.02	1.36 \pm 0.04	1.57 \pm 0.05
D	2.04 \pm 0.05	0.95 \pm 0.01	1.50 \pm 0.01	1.86 \pm 0.03
E	1.47 \pm 0.06	0.93 \pm 0.06	1.37 \pm 0.06	1.63 \pm 0.06
SCT	LA+QC	NLA	MLA	SLA
A	2.30 \pm 0.13	0.90 \pm 0.04	1.85 \pm 0.03	2.21 \pm 0.12
B	2.41 \pm 0.07	0.92 \pm 0.01	1.84 \pm 0.03	2.25 \pm 0.02
C	2.40 \pm 0.09	0.92 \pm 0.01	1.88 \pm 0.06	2.19 \pm 0.06
D	-	-	-	-
E	2.40 \pm 0.00*	0.90 \pm 0.00*	1.97 \pm 0.06	2.20 \pm 0.00*

*all replicate ratios were the same

Table 2: LA status of non-anticoagulated patient samples at Core and local/hospital laboratories: The number of positive and negative LA samples identified by both the laboratories and the ones that were discrepant are shown, as well as agreement and statistical results

		Core laboratories		Agreement	K Coeff	G-test*
		POS	NEG			
Local/hospital Laboratories	POS	98	5	115/132 87.1%	0.589 p<0.001	32.6 (99.8% confidence)
	NEG	12	17			

The G-test value as its over 99% confidence indicates that there is less than 0.2% chance of making a mistake regarding agreement of positive and negative LA results as defined.

Table 3: Analysis of possible reasons for discordant and equivocal results in LA status between Core and local/hospital laboratories testing in non-anticoagulated patient samples

	Discordant		Equivocal	
Total	17		34	
Local/hospital Laboratories	LA Pos	LA Neg	LA Pos	LA Neg
Samples tested on only one occasion at local/hospital laboratory	1	3	3	7
Tested more than once at local/hospital laboratory, but LA status varied	2	4	4	6
Method of LA assessment not specified or only one LA method performed	1	2	3	5
Tested more than once at local/hospital laboratory, with consistent results on LA status	1	3	3	3

Table 4: Agreement in LA status between Core and local/hospital laboratories results for anticoagulated patient samples: The number of positive and negative LA samples identified by both laboratories and the ones that were discrepant are shown, as well as agreement and statistical results

		Core laboratory		Agreement	K Coeff	G-test*
		POS	NEG			
Local/hospital Laboratories	POS	174	27	183/237	0.206	13.9
	NEG	27	9	77.2%	p<0.001	(99.98% confidence)

*The G-test value indicates that there is less than 0.02% chance of making a mistake regarding agreement of positive and negative LA results as defined in this study.

Table 5: Analysis of possible reasons for discordant and equivocal results in LA status between Core and local/hospital laboratories testing in anticoagulated patient samples

		Discordant		Equivocal	
Total		54		15	
Local/hospital laboratories		LA Pos	LA Neg	LA Pos	LA Neg
Samples tested on only one occasion at local/hospital laboratory		8	7	3	0
Tested more than once at local/hospital laboratory, but LA status varied		9	11	1	3
Method of LA assessment not specified or only one LA method		6	5	4	1





performed				
Tested more than once at local/hospital laboratory, with consistent results on LA status	4	4	2	1

Figure Legends





Figure 1: Dilute Russell's viper venom time (DRVVT) screen test used with Lupus anticoagulant (LA) positive (+) quality control (QC) plasma and for strong LA (SLA) plasma: All five Core laboratories used an ACL TOP500 analyser and performed the DRVVT with the same Lot numbers of HemosIL DRVVT Screen/Confirm. All labs were provided with a study protocol and laboratory manual for LA testing. Results are shown for clotting times and normalised ratios for the fresh vial of the 1st International Reference Panel strong positive (SLA) LA sample and the HemosIL LA Positive (LA+) QC, tested on each of three working days. The results were calculated as normalised ratios, using local pooled normal plasma, or where this was not available, a commercial pool or LA negative control plasma.

Figure 2: Silica clotting time (SCT) screen test used with LA+ QC plasma and for SLA plasma: All five Core laboratories used an ACL TOP500 analyser and performed the SCT with the same Lot numbers of HemosIL SCT Screen and SCT Confirm reagents. All labs were provided with a study protocol and laboratory manual for LA testing. Results are shown for clotting times and normalised ratios for the HemosIL LA Positive (LA+) QC plasma and the SLA sample from the 1st International Reference Panel for LA, tested on each of three working days. The results were calculated as normalised ratios, using local pooled normal plasma, or where this was not available, a commercial pool or LA negative control plasma.


Figure 3: Breakdown of non-anticoagulated and anticoagulated samples received from APS ACTION, samples excluded and final numbers included in analysis: Ninety patient samples (10 non-anticoagulated and 80 anticoagulated) were excluded from the final analysis due to: insufficient volume of citrate plasma for further analysis (n=38), or due to the local/hospital laboratory submitting only aCL/a β 2GPI results; or LA status not being available (n=52)



jth_14596_f1.tif



jth_14596_f2.tif



jth_14596_f3.tif