

Qin Qin (Orcid ID: 0000-0002-6432-2944)
Hernandez-Garcia Luis (Orcid ID: 0000-0003-3002-0304)
Liu Dapeng (Orcid ID: 0000-0002-4432-3202)
Nayak Krishna Shrinivas (Orcid ID: 0000-0001-5735-3550)
Schmid Sophie (Orcid ID: 0000-0003-0750-7798)
Van Osch Matthias JP (Orcid ID: 0000-0001-7034-8959)
Woods Joseph G. (Orcid ID: 0000-0002-0329-824X)
Zhao Moss Y. (Orcid ID: 0000-0002-0210-7739)
Zun Zungho (Orcid ID: 0000-0002-7297-3990)

Erratum to: Velocity-selective arterial spin labeling perfusion MRI: A review of the state of the art and recommendations for clinical implementation
(Magn Reson Med. 2022; 88:1528-1547)

Qin Qin^{1*}, David C. Alsop², Divya S. Bolar³, Luis Hernandez-Garcia⁴,
James Meakin⁵, Dapeng Liu¹, Krishna S. Nayak⁶, Sophie Schmid⁷,
Matthias J.P. van Osch⁷, Eric C. Wong³, Joseph G. Woods³, Greg Zaharchuk⁸,
Moss Y. Zhao⁸, Zungho Zun⁹, Jia Guo¹⁰,

*Corresponding Author: Qin Qin (PhD)
E-mail: qqin1@jhu.edu

In the above referenced article, we would like to correct the error of the misplaced SI_B term and update the Equation (1) as well as the first sentence of Section 9.5 as below:

9.5 Quantification of VSASL

The blood flow (BF, in mL/100 g/min) can be quantified based on the kinetic model as:

$$BF = \frac{6000 \cdot \lambda \cdot \Delta S \cdot e^{\frac{TI}{T_{1, \text{blood}}}}}{2 \cdot \alpha \cdot \tau \cdot SI_{PD} \cdot SI_B} \quad (1)$$

This is the author manuscript accepted for publication and has undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the Version of Record. Please cite this article as doi: 10.1002/mrm.29504

This article is protected by copyright. All rights reserved.